


Journal Pre-proof

Principles, potential and limitations of *ex vivo* basophil activation by flow cytometry in allergology: a narrative review.

Didier G. Ebo, MD PhD, Chris H. Bridts, MLT, Christel H. Mertens, MLT, Vito Sabato, MD PhD



PII: S0091-6749(20)31552-9

DOI: <https://doi.org/10.1016/j.jaci.2020.10.027>

Reference: YMAI 14817

To appear in: *Journal of Allergy and Clinical Immunology*

Received Date: 13 June 2020

Revised Date: 14 October 2020

Accepted Date: 20 October 2020

Please cite this article as: Ebo DG, Bridts CH, Mertens CH, Sabato V, Principles, potential and limitations of *ex vivo* basophil activation by flow cytometry in allergology: a narrative review., *Journal of Allergy and Clinical Immunology* (2020), doi: <https://doi.org/10.1016/j.jaci.2020.10.027>.

This is a PDF file of an article that has undergone enhancements after acceptance, such as the addition of a cover page and metadata, and formatting for readability, but it is not yet the definitive version of record. This version will undergo additional copyediting, typesetting and review before it is published in its final form, but we are providing this version to give early visibility of the article. Please note that, during the production process, errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

© 2020 Published by Elsevier Inc. on behalf of the American Academy of Allergy, Asthma & Immunology.

Principles, potential and limitations of *ex vivo* basophil activation by flow cytometry in allergology: a narrative review.

Didier G Ebo MD PhD ^{1,2}, Chris H Bridts MLT ¹, Christel H Mertens MLT ¹, Vito Sabato MD PhD ^{1,2}

¹ University of Antwerp, Faculty of Medicine and Health Sciences, Department of Immunology, Allergology, Rheumatology and the Infla-Med Centre of Excellence, Antwerp (Belgium) and Immunology, Allergology, Rheumatology, Antwerp University Hospital, Antwerp (Belgium)

³ Department of Immunology and Allergology, AZ Jan Palfijn Gent, Ghent, Belgium

***Correspondence:**

DG. Ebo MD PhD

University of Antwerp

Faculty of Medicine and Health Sciences

Immunology - Allergology – Rheumatology

Campus Drie Eiken T5.95

Universiteitsplein 1

2610 Antwerpen Belgium

Tel: ++ 32 (0) 3 2652595

immuno@uantwerpen.be

COI: The authors have no conflict of interest nor any financial relationship to declare.

Key words

Allergy, anaphylaxis, basophil, basophil activation, CD63, CD203c, flow cytometry, histamine, HistaFlow, IgE, mast cell, mast cell activation

Abbreviations

- 28 AIT: allergen specific immunotherapy
- 29 ASST: autologous serum skin test
- 30 BAT: basophil activation test
- 31 Bet v 1: major allergen from birch (*Betula verrucosa*) pollen
- 32 Can s: *Cannabis sativa*
- 33 CCD: cross-reactive carbohydrate determinant
- 34 CRD: component resolved diagnosis
- 35 DAO: diamino oxidase (histaminase)
- 36 DPT: drug provocation test
- 37 FCM: flow cytometry
- 38 FcεRI: high affinity receptor for IgE
- 39 fMLP: N-formylmethionyl-leucyl-phenylalanine
- 40 HDM: house dust mite
- 41 Hev b: *Hevea brasiliensis*
- 42 HVA : hymenoptera venom allergy
- 43 IDHR: immediate drug hypersensitivity reaction
- 44 IL: interleukin
- 45 LTP : lipid transfer protein
- 46 Mal d 1 : Bet v 1 homologue in apple
- 47 MC: mast cell
- 48 MRGPRX2: Mas-related G protein coupled receptor X2
- 49 OEEP2: oxygen-evolving enhancer protein 2
- 50 P38 MAPK: p38 mitogen-activated protein kinase
- 51 sIgE: specific IgE
- 52 SLIT: sublingual immunotherapy
- 53 SPT: skin prick test
- 54 STAT5: signal transducer and activator of transcription 5
- 55 STs: skin tests
- 56 TG-ROC: two-graph receiver-operating characteristic

57 VIT: venom immunotherapy

58

59

60

Journal Pre-proof

Abstract

The major challenge of allergy diagnosis lies in the development of accessible and reliable diagnostics allowing correct prediction of the clinical outcome upon exposure to the offending allergen(s) and cross-reactive structures. Since the late nineties, evidence has accumulated that flow-assisted analysis and quantification of *ex vivo* activated basophils (BAT) might meet this requirement for different IgE-dependent allergies and particular forms of auto-immune urticaria. Other, so called non-diagnostic applications of the BAT involve therapeutic monitoring, follow-up of natural histories and identification of allergenic recognition sites. However, it has also become clear that appropriate use of the BAT necessitates knowledge about degranulation metrics and guidance to guarantee correct execution and interpretation of the results. Here, we review the most relevant applications and limitations of the BAT. Some personal statements and views about its perspectives are made.

Introduction

Degranulation of basophils and mast cells (MCs) can be triggered by various processes (Figure 1). IgE-dependent degranulation involves synthesis of allergen-specific IgE antibodies (sIgE) by plasma cells. These sIgE antibodies bind to their high affinity receptors (FcεRI) present on the surface membrane of MCs and basophils to form sIgE/FcεRI complexes. Encounter of specific allergen that cross-links such sIgE/FcεRI complexes can induce degranulation with release of preformed mediators and *de novo* synthesis of inflammatory mediators. However, IgE-mediated activation is not only achieved by traditional allergens, it can also occur by lectins with a binding specificity that matches the glycosylation of IgE and/or the FcεRI, or by other molecules such as superallergens (protein Fv, HIV gp120) or *S. mansoni* IPSE/alpha-1^{1,2}. IgE-independent activation results from occupation of various receptors, e.g. C3aR and C5aR (receptors for the anaphylatoxins C3a and C5a) or the Mas-related G protein-coupled receptor MRGPRX2 on MCs by drugs, as observed in some immediate drug hypersensitivity reactions (IDHRs)^{3,4}.

Generally, diagnosis of sIgE-mediated allergies starts with a history taking together with skin tests (STs) and/or quantification of sIgE (including component resolved diagnosis, CRD). However, as none of these tests is absolutely predictive, many have focused on *ex vivo* basophil activation tests (BATs) to close the gaps in their diagnostic instrumentation and to avoid sometimes potentially dangerous provocations. In BATs, functionality of the cells can be explored via quantification of released mediators and phenotyping of intracellular and/or surface changes by flow cytometry (FCM). This review focuses on the main applications and limitations of FCM-based BATs and provides some guidance to guarantee correct execution and interpretation of these tests. It appears that BATs are more than a diagnostic aid but that further standardization and harmonization is required before its entrance in mainstream use. Most of the points to consider relate to the infrastructure and expertise of the laboratory, choice of read-out, allergen preparation, metrics of basophil activation, responder status of the cells and correct positioning of the BAT in the diagnostic algorithm. Needless to stress that state-of-the-art equipment and expertise is required for correct execution of commercial and/or homemade BATs and that inter-laboratory comparison of results should benefit from harmonization and standardization of both instruments and protocols. Users of BATs must realize that data from different laboratories are not always readily interchangeable and that local validation procedures are mandatory.

The basophil activation test via flow cytometry: principles and technical aspects

Basophil identification and activation/degranulation markers

The foundations of modern FCM-based BATs date from 1991, with the first description of the lysosomal associated membrane protein CD63 to be a degranulation marker of basophils⁵. At present, different protocols allowing detection of surface marker alterations, intracellular changes and exteriorization of granular content have been developed⁶⁻⁸. As shown in Figures 1 and 2 and Table E1 of repository, traditional FCM-based BATs use whole blood and rely upon cellular identification and quantification of activation and/or degranulation markers on the surface membrane. These changes are detectable and quantifiable on an individual cell level using specific fluorescent-labelled monoclonal antibodies. Actually, in most studies, basophils have been identified by scatter characteristics, reflecting cellular size and granularity combined with staining of surface markers such as IgE/CD203c, CCR3/CD3 or CD123/HLA-DR. Subsequently, after activation, the appearance or up-regulation of specific markers, such as lysosomal CD63 or the lineage-specific ectoenzyme CD203c is measured, with degranulating basophils being defined as CD203c⁺⁺CD63⁺ cells. Based upon divergent time-kinetics, responses to secretagogues/inhibitory compounds, signalosome and absence/presence of mediator release, a CD63- and CD203c-compartment have been identified. The “CD203c-compartment” is characterized by a rapid and significant upregulation of CD11b, CD13, CD164 and CD203c⁹⁻¹¹. In the “CD63-compartment”, maximum upregulation of CD63 and CD107a is slower and reflects anaphylactic degranulation¹¹. For a comprehensive description of these activation marker profiles, their different relationship to piecemeal and anaphylactic degranulation and their signal transduction processes the reader is referred elsewhere in this Journal¹². Whether the CD63- and/or CD203c-based microfluidic immunoaffinity BAT technique¹³, and mass cytometry (CyTOF)¹⁴ can keep promise remains to be established. Other surface markers that can be used to quantify *ex vivo* basophil activation and degranulation are the inhibitory receptors CD300a and CD200R and the activating receptor CD300c¹⁵⁻¹⁷. Besides, basophilic activation can also be analysed by studying the phosphorylation status of signalling molecules such as p38 mitogen-activated protein kinase (MAPK) or signal transducer and activator of transcription (STAT)5^{18, 19}. Degranulation of basophils can also be explored by measuring the exteriorization of granule matrix and decrease in intracellular histamine content. Briefly, anionic proteoglycans from exteriorized basophil granule matrix are stained by cationic fluorescent avidin probes^{6, 20, 21}, whereas the quantification of the intracellular histamine relies upon a histaminase affinity technique in which diamino oxidase is coupled to a fluorochrome²². Newly synthesized cytokines such as interleukin (IL) 4 and IL-13, or their mRNA, can be trapped and measured intracellularly²³. Finally, basophil activation can also be measured by imaging of intracellular Ca⁺⁺²⁴.

Whole blood versus separated cells and effects of IL-3

Most applied BAT methods use whole blood and study cells in their “natural” environment. However, there are various strategies to enrich and/or purify the cells²⁵. If the analyses are performed in whole blood, one should verify the possibility for *in vivo* stimulation, e.g. by exposure to IL-3. Such exposure is not associated with an up-regulation of CD63, but might be demonstrable by an up-regulated expression of CD69, CD203c, p38MAK and STAT5^{6, 18, 19, 26, 27}. Note that in whole blood settings the outcome of the test can be influenced by blocking antibodies (through CD32)²⁸ or stoichiometrically via interference with the IgE-allergen interaction²⁹. These blocking antibodies can be deleted by washing leukocytes or using purified basophils²⁵. Although modern purification protocols are designed to avoid cell loss and activation, e.g. by using low temperatures, Ca²⁺/Mg²⁺-free buffers (with added EDTA), and using negative immunomagnetic selection, cell loss and background activation can occur^{30, 31}. Alternatively, basophils can be primed intentionally with IL-3 to optimize analytical sensitivity, that is, responsiveness to allergen^{6, 32, 33} and autoreactive antibodies³⁴.

Passive sensitization experiments

Although protocols have been developed to permit BATs up to 24 hours after sampling^{14, 35}, a major weakness of the test remains the necessity for viable basophils, as sensitivity of the test decreases over time^{33, 36}. To circumvent this limitation, several groups have adopted BATs with passive sensitization of stripped donor basophils^{37, 38}. This involves stripping of bound sIgE antibodies from their surface FcεRI receptors with the aid of acidic buffers, incubation of these stripped cells with patient’s serum (containing sIgE antibodies to the allergen being investigated) and finally challenging the passively sensitized basophils with the allergen at the first stage of the BAT. Donor cells for sensitizing should be from a healthy subject whose basophils are known to be good responders. Both unstripped and stripped donor basophils should be included in the controls. This procedure, apart from being laborious with a number of extra steps, carries the risk of non-specific stimulation or damage to basophils and is difficult to standardize. Results so far indicate rather varying performance. For example, in the study by Moneret-Vautrin et al³⁹, a food-specific IgE results between of 3.50 and 35 kUA/L was required for effective passive sensitization. In the study by Mueller-Wirth et al⁴⁰, passive sensitization was already demonstrable at drug-specific IgE titers of 1.0 KUA/L. Whether, cell lines (e.g. rat basophil leukemia (RBL) or LAD2 MCs) or donor MCs constitute valuable alternatives to circumvent the limitations of BAT remains to be established, but preliminary results seem promising⁴¹⁻⁴⁴. In the comparative study of Larsen et al³⁷, BAT was shown to be more performant than histamine release testing and passive histamine release testing.

Selection of the optimal “allergen” and “dose” (degranulation metrics)

Other crucial elements to keep in mind are correct selection of source material, preparation and storage of the allergen extract, and metrics of the allergen dose-response curves¹². For many applications, BATs use crude allergen extracts⁴⁵⁻⁴⁷. However, the inherent variability and instability of natural allergens, complicate selection of the best source material and optimal extraction procedure (for review see: ⁴⁸). For example, thermal processing can influence the capacity of peanuts to trigger basophils⁴⁹. This impact seems highly divergent between patients and unpredictable by SDS-PAGE or IgE binding⁴⁹. Or, as shown in Figure E1 of the repository file, in contrast to sesame oil, a whole sesame seed extract might not trigger basophil degranulation in a patient who experienced sesame oil anaphylaxis. Fortunately, it is not all doom and gloom. On several occasions, BATs have shown to benefit diagnosis in difficult patients demonstrating clinically irrelevant sIgE results because of sensitization to cross-reactive carbohydrate determinants (CCDs)^{50, 51}. For a summary on allergen concentrations see ⁵². Note, these concentrations are only indicative and might not apply to laboratories using different equipment and protocols. It is also important to remember that there can be a significant difference between the stock concentration and final concentration in the aliquot. For drugs and related compounds, it is recommended to express concentrations on a molar base.

Basophil responses are characterized by a broad variability necessitating use of different stimulation concentrations enabling construction of dose-finding curves (Figure 3). These curves encompass different metrics including: basophil sensitivity, EC50 and basophil reactivity that differ according to the stimulation conditions and applied read-out. For an explanation and implications of the metrics of dose-response curves, the reader is referred elsewhere^{12, 53-56}. Clearly, clinical validation of BATs cannot be considered appropriate when it failed to carefully establish allergen-specific dose-responses. Ideally, these dose-responses show a sigmoidal shape with plateau or bell-shape. However, because of complexity of most allergens and relative affinity of different epitope-paratope interactions, dose-response curves can show unpredictable complex courses that can vary among allergens and subjects. In diagnostic settings, where only a limited number of allergen concentrations are employed, there is a chance of producing false-negative results. Nevertheless, in our experience, it is not rare to find one or two optimal specific stimulation concentrations, even for drugs that can induce false-negative results because of cytotoxicity. Other explanations for false-negative results are: a non-responders status of the cells (as is observed in about 5-15% of the patients), poor storage conditions of the blood sample (analyses are best performed within 4 hours³³), use of cytotoxic (concentrations of) allergens, degraded allergens⁵⁷, (pharmacologic) interference with surface receptors⁵⁸, inhibition by cross-reactive compounds⁵⁹, and blunted responses because of a preactivated status of the cell³⁰. Therefore, BATs should always include a negative control setting to assess spontaneous expression of the readout, a positive control to verify responsiveness of the cell

upon cross-linking of surface sIgE/FcεRI complexes. fMLP, that acts independent from IgE/FcεRI can be of benefit to include such a positive control to verify cell viability. In “non-responders”, who do not react to positive control and allergen, negative results should be considered as uninterpretable. If cells do not respond to positive control but do to allergen, the test can be considered as positive, provided there is no nonspecific stimulation in at least 3-5 (exposed) control individuals. Non-responsiveness is attributed to disturbances in the signalosome of the FcεRI-pathway, particularly failure to express the downstream tyrosine kinase Syk. Although IL-3 can restore non-releaser status, this approach is of little help in traditional BATs as it takes days for conversion⁶⁰. It should be kept in mind that IL-3 can blunt responses measured via upregulation of expression of CD69 and CD203c and phosphorylation of p38MAPK and STAT5.

Determination of the decision threshold (cut-off limit)

Validation of a diagnostic cannot be considered appropriate when it failed to establish a decision threshold differentiating between positive reactions and responses in control individuals. In such studies one should establish optimal allergen-specific decision thresholds and abandon pre-defined arbitrarily chosen cut-off limits for determination of sensitivity, specificity, predictive values and performance of the test. Normally, the cutoff of in vitro tests is defined by the mean of blank tests + 3.3 SD. However, for BAT this definition is rarely applied. As shown in Figure 3, an alternative method to calculate optimal decision thresholds is analyses of two-graph receiver-operating characteristic (TG-ROC) curves. In TG-ROC curves, the test sensitivity and specificity are plotted against the threshold limit assuming the latter to be an independent variable. For rare conditions it might be difficult to include a minimum of patients to construct TG-ROC curves. In such cases, comparison of the results in the patient(s) with control experiments in minimally 3-5 (exposed) control individuals is proposed. The importance of cut-offs in BAT is stressed by Dreborg, who argues the use of poorly documented decision thresholds⁶¹.

Clinical applications

Over 25 years, BATs have evolved into a useful test in the evaluation of patients with inhalant, food, Hymenoptera venom, *Hevea* latex, immediate drug allergy and some forms of chronic urticaria/angioedema. However, its place within the diagnostic algorithms is highly variable and often still poorly determined. Actually, the position of the BAT is determined by its ease to ascertain a correct execution and interpretation, and the availability of (a) safe, more accessible, better performing alternative(s). For example, as drugs are manufactured according to GMP principles, easily accessible, and only few alternative *in vitro* tests are available, it is likely BATs to deserve a more predominant place in the diagnostic algorithm for IDHRs than for other IgE-mediated allergies. For inhalant, food, Hymenoptera venom and *Hevea* latex allergy, extract preparation and

interpretation of results might be less obvious and the merit of the test must be seen in a more global context of other available diagnostics, including CRD.

Inhalant allergy

As shown in Table E2 of the repository file, utility of the BAT has been explored in allergy to house dust mite (HDM), pollen and cat using natural extracts and purified/recombinant components. Although overall performance of BATs in inhalant allergy is good, the technique is rarely beneficial. Diagnosis of inhalant allergies can readily be established by other means such as STs and measurement of sIgE, including CRD. However, in cases of “entopy”, that is, local allergic rhinitis, with positive nasal provocation tests and negative SPTs and sIgE, BATs allowed diagnosis in 8 out of 16 patients⁶². Correlations between basophil sensitivity and nasal/bronchial provocation tests as well as asthma severity and efficacy of omalizumab treatment have been described^{53, 63}. BATs have also been used to monitor allergen-specific immunotherapy (AIT) for, HDM, birch, timothy grass, *Lolium perenne*, mugwort, *Parietaria*, Japanese cedar and cypress⁶⁴⁻⁶⁸. In these studies, reduced basophil sensitivity occurs early during AIT and is likely due to interference of blocking IgG antibodies.

Food allergy

The performance of the BAT in food allergy has already extensively been studied and reviewed. From these reports, the BAT has emerged as a potential diagnostic for primary and secondary food allergies including the tick-borne alpha-gal syndrome⁶⁹⁻⁷¹. However, it is premature to claim BATs to be “food challenges in a test tube”⁷⁰. The utility of BATs is allergen-dependent and requires validation for different allergens and phenotypes (e.g. oral allergy syndrome vs. anaphylaxis). However, such a validation might be challenging for reasons already addressed above but also because of distinct age and geographic-related sensitization patterns^{72, 73}. Finally, it should be kept in mind that the test performance can depend on the control group. For example, in exploring its performance in the context of cross-reactivity syndromes, a comparison between patients and healthy control individuals is likely to result in an overestimation of its specificity^{51, 73, 74}. In other words, a comparison between true patients and healthy control individuals might not be representative for the general clinic population in which one might need to differentiate between clinically relevant and irrelevant sensitization, rather than to dichotomize between patients and controls. Table E3 of the repository file summarizes the sensitivity and specificity from BATs in food allergy. Predictive values are summarized in table E9. BATs can be useful to discriminate between clinically relevant and irrelevant sIgE results or when no alternative *in vitro* tests are available^{55, 74-76} or for reducing the need for challenges in difficult cases who experienced severe anaphylaxis^{70, 77}. Alternatively, as already mentioned higher, one should keep in mind the possibility of clinically irrelevant BAT results induced by dietary lectins^{1, 2}.

With respect to BATs using components, it is noteworthy a single component rarely to cover the entire sensitization profile^{78, 79}, and that outcomes have been reported to be age and/or population dependent⁷³. BATs have also been used to monitor AIT and allergen-specific oral immunotherapy (OIT) for peanut^{56, 80-82}, sesame⁸³, cow's milk⁸⁴⁻⁸⁶ and egg⁸⁷. In line with the findings for aeroallergens, reduced basophil allergen sensitivity during AIT and OIT food is likely due to interference of blocking sIgG antibodies^{80-83, 86, 87}. Acosta *et al*⁸⁸, recently provided an immunological explanation why birch-associated apple allergy cannot be effectively treated by administration of the sensitizing pollen allergen (SLIT with recombinant (r) Bet v 1, the major allergen from birch (*Betula verrucosa*)). They found that treatment with rBet v 1 promotes specific IgG antibodies that cross-react with rMal d 1 from apple (*Malus domestica*) but lack sufficient affinity to inhibit BAT with apple allergens and to induce cross-protection. In contrast, treatment with the apple allergen induced a food allergen-specific de novo antibody response characterized by IgG1 antibodies with IgE-blocking bioactivity and specific for epitopes exclusive of the apple allergen. Some have claimed BATs also to predict clinical severity and prognosis of food allergy^{54, 89-91} and food challenge responses^{55, 91-93}, thresholds of reactivity^{54, 94}, to help determination when food can safely be (re)introduced⁹⁵, and to distinguish degrees of tolerance⁹⁶. However, not all studies seem promising^{72, 97}. Therefore, and because of the possibility of reporting bias, additional comprehensive studies are required for confirmation. Ideally, these studies should compare the BAT with STs and/or CRD, as these might provide similar information but in an easier manner. Recently passive BAT was used to demonstrate that mammalian glycolipid can activate allergic effector cells via surface bound sIgE in alpha-gal allergy³⁸ and to identify unique epitopes of certain peanut components⁹⁸. BAT has also been used to monitor the effect of omalizumab in food allergy⁹⁹.

Hymenoptera venom allergy (HVA)

The utility of the BAT in wasp and honeybee venom allergy has been largely explored (Table E4 of repository). BAT can benefit diagnosis of HVA, especially in difficult cases that yield equivocal or negative sIgE and/or ST results. About 4-6% of patients with HVA demonstrate negative sIgE and ST results. In some of these cases, BATs can be useful to identify the culprit insect and guide treatment^{100, 101}. Besides, BATs can also help to take the sting out of difficult cases presenting with double positive sIgE results resulting from sensitization to α -1,3-fucose containing CCDs present on various Hymenoptera venom proteins^{100, 102, 103}. However, utility of the BAT in HVA needs to be reviewed in the global the context of other diagnostics. With the venue of CRD using non-glycosylated recombinant proteins, the BAT likely lost ground but the technique remains useful in patients with negative skin test and sIgE investigations. BAT appears not predictive for severity of sting reactions^{104, 105}. With respect to venom immunotherapy (VIT) it has been shown treatment to decrease basophil sensitivity but not reactivity^{104, 106-110}. Effects of VIT on basophils include early basopenia

and a decrease in intracellular histamine content during maintenance treatment¹⁰⁹. Some studies suggest basophil sensitivity to be predictive for side effects during the build-up phase of VIT^{107, 111}. However, this is not the experience of others¹¹². In patients with mastocytosis the BAT adds little, if at all, to the diagnosis in cases with negative sIgE and ST results^{113, 114}.

Hevea latex allergy

As shown in Table E5 of the supplementary file, the first descriptions of BAT in the diagnosis of allergy to latex from *Hevea brasiliensis* dates back from almost 2 decades ago⁵¹. The technique predominantly proved to be helpful to discriminate between clinically relevant and irrelevant sIgE results, the latter mainly resulting from sensitisation to CCD and profilin¹¹⁵. However, as in HVA, with the venue of CRD using non-glycosylated components, utility of the test lost ground. Alternatively, the BAT can help when other tests are unavailable, e.g. sensitization to Hev b 12, the non-specific LTP from *Hevea*¹¹⁶.

Cannabis sativa

Allergy to *Cannabis sativa* (Can s) has become a significant health problem and can be associated with complex cross-reactivity syndromes involving many vegetables, fruits and latex^{117, 118}. Diagnosis of cannabis allergy can be challenging, mainly because of the moderate specificity of sIgE, STs or BATs using natural extracts¹¹⁹. In such difficult cases, BAT with Can s 3, the non-specific lipid protein of *C. sativa* can benefit correct diagnosis; especially as traditional sIgE testing is still not available¹¹⁹. The utility of BAT with Can s 4, the oxygen-evolving enhancer protein 2 (OEEP2), in Can s 3 negative patients, remains to be established¹²⁰. Note that Can s 3 covers approximately two-thirds of the Cannabis IgE reactivity profile^{118, 119}.

Drug hypersensitivity

Although drug provocation tests (DPT) are considered the gold standard, DPT might be difficult because of ethical and practical considerations. By consequence, confirmation of IDHRs generally relies on STs and quantification of sIgE. Unfortunately, diagnosis of IDHRs is not always straightforward, mainly because of uncertainties associated with STs and unavailability of drug-sIgE assays. As shown in Tables E6-E8 from the repository file, many have explored BATs as a confirmatory diagnostic in IDHRs. It has emerged the performance of BATs in IDHRs to vary considerably according to the investigated drug (class), decision threshold^{61, 121}, time elapsed between index reaction and testing¹²². Alternatively, application of *ex vivo* basophil experiments in IDHRs might extend beyond diagnosis¹²³. Different groups have claimed utility of the BAT in diagnosing IDHRs to neuromuscular blocking agents, β -lactam antibiotics, carboplatin, opiates, iodinated contrast media, biologicals¹²⁴⁻¹²⁶. However, for some of these drug(s) (classes) evidence is limited¹²⁷ or controversial^{57, 121, 128, 129}. In our experience, BAT adds little to the diagnosis of (amino)penicillin hypersensitivity (unpublished data). In addition, the technique could deepen our

insights in immune IgE/FcεRI and non-immune mechanisms of IDHRs^{59, 127, 129, 130}, unveil cross-reactivity between structurally related substances^{127, 131, 132}, benefit identification of antibody recognition sites¹³⁰ and monitor effects of rapid desensitisation strategies^{133, 134}. For some drugs such as opiates, BAT is likely to constitute the only diagnostic^{127, 135}. Figure E2 of the repository file shows an example of a BAT to opiates in a patient who experienced anaphylaxis from pholcodine but tolerated a codeine and morphine DPT.

It is evident that the BAT only adds to diagnosis in IDHRs that involve basophil degranulation. BATs cannot document IDHR resulting from enzymatic inhibition of cyclo-oxygenase as happens in non-selective hypersensitivity to non-steroidal anti-inflammatory drugs (NSAIDs) and angioedema to angiotensin converting enzyme inhibitors. Therefore, we do not advocate use of the BAT in this context. Alternatively, for selective NSAID reactors, BAT might benefit diagnosis and help to depict presence or absence of parent drug/metabolite reactive IgE antibodies¹³⁶.

For diagnostic purposes, the fact that positive basophil responses cannot discriminate between sIgE/FcεRI cross-linking and alternative mechanisms is not a drawback. However, evidence of usefulness of BAT in IDHRs from IgE/FcεRI-independent pathways is limited and should be interpreted carefully. The reporting, basophils to constitutively express the MRGPRX2 receptor³⁰ contradicts earlier observations^{8, 137} and is difficult to align with the observation various MRGPRX2 agonists not to trigger basophil degranulation in control individuals^{121, 128, 129}. Figure E3 of the supplementary file shows plots of MRGPRX2 by MCs and basophils⁸. Figure 4 shows an algorithm indicating the place of BAT in the diagnostic work-up of IDHR. Particular attention is paid to its place in discriminating between sIgE/FcεRI- and MRGPRX2-mediated reactions.

Urticaria and angioedema

Acute and chronic histaminergic urticaria's and angioedema's rest upon MC and basophil degranulation via diverse innate and adaptive immune responses, including auto-immune processes^{138, 139}. At present, in chronic spontaneous urticaria, two groups of MC degranulating signals have been identified, that is IgE autoantibodies to auto-allergens and IgG autoantibodies that target FcεRI or IgE/FcεRI complexes present on the MC surface. The presence of such anti-IgE or anti-FcεRI antibodies can be assessed functionally using patients' sera in an autologous serum skin test (ASST) and/or to passively sensitize donor basophils in the autoimmune BAT (Figure E4 of the repository file)¹⁴⁰⁻¹⁴².

Conclusion and future directions

The BAT provides the physician and clinical laboratory with a promising diagnostic approach, especially in difficult cases where traditional tests are unavailable or yield uncertain results. However, before entrance in mainstream application, additional larger scale studies are required to

confirm and critically verify some observations. Ideally, such studies should assess how the BAT relates to other diagnostics. Further automation of data analyses and bioinformatic tools should advance standardization and quality assurance and thus accelerate transition to the clinics¹⁴³. Although the non-diagnostic applications of BAT are still in their infancy, with increasing employment, it is expected the technique to become an attractive and valuable asset to study various domains of basophil activation/degranulation biology. Evidence is emerging that the BAT might help to deepen our understandings in mechanistic endotypes of IDHRs, benefit identification of antibody recognition sites, expand our understandings of desensitisation and tolerance induction strategies, predict natural disease courses and prognosis.

Acknowledgements

Didier Ebo is a senior clinical researcher of the Research Foundation Flanders (FWO: 1800614N). Vito Sabato is a senior clinical researcher of the Research Foundation Flanders (FWO: 1804518N). The authors would like to thank Prof. Dr. Jean-Pierre Timmermans and Dr. Isabel Pintelon of the Antwerp Centre of Advanced Imaging (ACAM), Department of Veterinary Sciences, University of Antwerp, Antwerp (Belgium) for the use of the Leica SP8/LSM confocal microscope (GOH4216N).

400

401

402

Journal Pre-proof

Journal Pre-proof

1. Pramod SN, Venkatesh YP, Mahesh PA. Potato lectin activates basophils and mast cells of atopic subjects by its interaction with core chitobiose of cell-bound non-specific immunoglobulin E. *Clin Exp Immunol* 2007; 148:391-401.
2. Krithika N, Pramod SN, Mahesh PA, Venkatesh YP. Banana lectin (BanLec) induces non-specific activation of basophils and mast cells in atopic subjects. *Eur Ann Allergy Clin Immunol* 2018; 50:243-53.
3. Finkelman FD, Khodoun MV, Strait R. Human IgE-independent systemic anaphylaxis. *J Allergy Clin Immunol* 2016; 137:1674-80.
4. McNeil BD, Pundir P, Meeker S, Han L, Undem BJ, Kulka M, et al. Identification of a mast-cell-specific receptor crucial for pseudo-allergic drug reactions. *Nature* 2015; 519:237-41.
5. Knol EF, Mul FP, Jansen H, Calafat J, Roos D. Monitoring human basophil activation via CD63 monoclonal antibody 435. *J Allergy Clin Immunol* 1991; 88:328-38.
6. Ebo DG, Elst J, van Houdt M, Pintelon I, Timmermans JP, Horiuchi T, et al. Flow cytometric basophil activation tests: Staining of exteriorized basophil granule matrix by fluorescent avidin versus appearance of CD63. *Cytometry B Clin Cytom* 2020. doi:10.1002/cyto.b.21868.
7. Bridts CH, Sabato V, Mertens CH, Hagendorens MM, De Clerck LS, Ebo DG. Flow Cytometric Allergy Diagnosis: Basophil Activation Techniques. *Methods Mol Biol.* 2020;2163:183-195. doi: 10.1007/978-1-0716-0696-4_15.PMID: 32766976.
8. Elst J, Sabato V, Hagendorens MM, Van Houdt M, Faber MA, Bridts CH, et al. Measurement and Functional Analysis of the Mas-Related G Protein-Coupled Receptor MRGPRX2 on Human Mast Cells and Basophils. *Methods Mol Biol.* 2020;2163:219-226. doi: 10.1007/978-1-0716-0696-4_18.PMID: 32766979.
9. Hennersdorf F, Florian S, Jakob A, Baumgartner K, Sonneck K, Nordheim A, et al. Identification of CD13, CD107a, and CD164 as novel basophil-activation markers and dissection of two response patterns in time kinetics of IgE-dependent upregulation. *Cell Res* 2005; 15:325-35.
10. MacGlashan D, Jr. Marked differences in the signaling requirements for expression of CD203c and CD11b versus CD63 expression and histamine release in human basophils. *Int Arch Allergy Immunol* 2012; 159:243-52.
11. MacGlashan D, Jr. Expression of CD203c and CD63 in human basophils: relationship to differential regulation of piecemeal and anaphylactic degranulation processes. *Clin Exp Allergy* 2010; 40:1365-77.
12. MacGlashan DW, Jr. Basophil activation testing. *J Allergy Clin Immunol* 2013; 132:777-87.
13. Aljadi Z, Kalm F, Nilsson C, Winqvist O, Russom A, Lundahl J, et al. A novel tool for clinical diagnosis of allergy operating a microfluidic immunoaffinity basophil activation test technique. *Clin Immunol* 2019; 209:108268.
14. Mukai K, Gaudenzio N, Gupta S, Vivanco N, Bendall SC, Maecker HT, et al. Assessing basophil activation by using flow cytometry and mass cytometry in blood stored 24 hours before analysis. *J Allergy Clin Immunol* 2017; 139:889-99.e11.
15. Gibbs BF, Sabato V, Bridts CH, Ebo DG, Ben-Zimra M, Levi-Schaffer F. Expressions and inhibitory functions of CD300a receptors on purified human basophils. *Exp Dermatol* 2012; 21:884-6.
16. Sabato V, Verweij MM, Bridts CH, Levi-Schaffer F, Gibbs BF, De Clerck LS, et al. CD300a is expressed on human basophils and seems to inhibit IgE/FcepsilonRI-dependent anaphylactic degranulation. *Cytometry B Clin Cytom* 2012; 82:132-8.
17. Zenarruzabeitia O, Vitale J, Terren I, Orrantia A, Astigarraga I, Dopazo L, et al. CD300c costimulates IgE-mediated basophil activation, and its expression is increased in patients with cow's milk allergy. *J Allergy Clin Immunol* 2019; 143:700-11.e5.
18. Ebo DG, Dombrecht EJ, Bridts CH, Aerts NE, de Clerck LS, Stevens WJ. Combined analysis of intracellular signalling and immunophenotype of human peripheral blood basophils by flow cytometry: a proof of concept. *Clin Exp Allergy* 2007; 37:1668-75.

19. Verweij MM, Sabato V, Nullens S, Bridts CH, De Clerck LS, Stevens WJ, et al. STAT5 in human basophils: IL-3 is required for its FcεRI-mediated phosphorylation. *Cytometry B Clin Cytom* 2012; 82:101-6.
20. Joulia R, Mailhol C, Valitutti S, Didier A, Espinosa E. Direct monitoring of basophil degranulation by using avidin-based probes. *J Allergy Clin Immunol* 2017; 140:1159-62.e6.
21. Mukai K, Chinthrajah RS, Nadeau KC, Tsai M, Gaudenzio N, Galli SJ. A new fluorescent-avidin-based method for quantifying basophil activation in whole blood. *J Allergy Clin Immunol* 2017; 140:1202-6.e3.
22. Ebo DG, Bridts CH, Mertens CH, Hagendorens MM, Stevens WJ, De Clerck LS. Analyzing histamine release by flow cytometry (HistaFlow): a novel instrument to study the degranulation patterns of basophils. *J Immunol Methods* 2012; 375:30-8.
23. Devouassoux G, Foster B, Scott LM, Metcalfe DD, Prussin C. Frequency and characterization of antigen-specific IL-4- and IL-13- producing basophils and T cells in peripheral blood of healthy and asthmatic subjects. *J Allergy Clin Immunol* 1999; 104:811-9.
24. Heinemann A, Ofner M, Amann R, Peskar BA. A novel assay to measure the calcium flux in human basophils: effects of chemokines and nerve growth factor. *Pharmacology* 2003; 67:49-54.
25. Falcone FH, Gibbs BF. Purification of Basophils from Peripheral Human Blood. *Methods Mol Biol* 2020;2163:35-48. doi: 10.1007/978-1-0716-0696-4_2.
26. Yoshimura C, Yamaguchi M, Iikura M, Izumi S, Kudo K, Nagase H, et al. Activation markers of human basophils: CD69 expression is strongly and preferentially induced by IL-3. *J Allergy Clin Immunol* 2002; 109:817-23.
27. Hauswirth AW, Sonneck K, Florian S, Krauth MT, Böhm A, Sperr WR, et al. Interleukin-3 promotes the expression of E-NPP3/CD203C on human blood basophils in healthy subjects and in patients with birch pollen allergy. *Int J Immunopathol Pharmacol* 2007; 20:267-78.
28. Cady CT, Powell MS, Harbeck RJ, Giclas PC, Murphy JR, Katial RK, et al. IgG antibodies produced during subcutaneous allergen immunotherapy mediate inhibition of basophil activation via a mechanism involving both FcγRIIA and FcγRIIB. *Immunol Lett* 2010; 130:57-65.
29. Ejrnaes AM, Svenson M, Lund G, Larsen JN, Jacobi H. Inhibition of rBet v 1-induced basophil histamine release with specific immunotherapy -induced serum immunoglobulin G: no evidence that FcγRIIB signalling is important. *Clin Exp Allergy* 2006; 36:273-82.
30. Wedi B, Gehring M, Kapp A. The pseudoallergen receptor MRGPRX2 on peripheral blood basophils and eosinophils: Expression and function. *Allergy*. 2020 Sep;75:2229-2242. doi: 10.1111/all.14213. Epub 2020 Feb 15. PMID: 32003863. Comments in: Surface expression of MRGPRX2 on resting basophils: An area of controversy. Sabato V, Elst J, Van Houdt M, Bridts C, Mertens C, Ebo DG. *Allergy*. 2020 Sep;75:2421-2422. doi: 10.1111/all.14252. PMID: 32929729 and Reply to Sabato V et al "Surface expression of MRGPRX2 expression on resting basophils: An area of controversy". Wedi B, Gehring M, Kapp A. *Allergy*. 2020 Sep;75:2424-2427. doi: 10.1111/all.14446.
31. Falcone FH, Gibbs BF. Purification of basophils from peripheral human blood. *Methods Mol Biol* 2014; 1192:35-47.
32. Black KM, Lussier AM, Gion WR, Kasaian MT. Cytokine priming of human basophils: description of allergen 'nonreleasers'. *Int Arch Allergy Immunol* 1996; 111:142-51.
33. Sturm EM, Kranzelbinder B, Heinemann A, Groselj-Strele A, Aberer W, Sturm GJ. CD203c-based basophil activation test in allergy diagnosis: characteristics and differences to CD63 upregulation. *Cytometry B Clin Cytom* 2010; 78:308-18.
34. Gentinetta T, Pecaric-Petkovic T, Wan D, Falcone FH, Dahinden CA, Pichler WJ, et al. Individual IL-3 priming is crucial for consistent in vitro activation of donor basophils in patients with chronic urticaria. *J Allergy Clin Immunol* 2011; 128:1227-34.e5.

35. Sanz ML, Sanchez G, Gamboa PM, Vila L, Uasuf C, Chazot M, et al. Allergen-induced basophil activation: CD63 cell expression detected by flow cytometry in patients allergic to *Dermatophagoides pteronyssinus* and *Lolium perenne*. *Clin Exp Allergy* 2001; 31:1007-13.
36. Sturm GJ, Kranzelbinder B, Sturm EM, Heinemann A, Groselj-Strele A, Aberer W. The basophil activation test in the diagnosis of allergy: technical issues and critical factors. *Allergy* 2009; 64:1319-26.
37. Larsen LF, Juel-Berg N, Hansen KS, Clare Mills EN, van Ree R, Poulsen LK, et al. A comparative study on basophil activation test, histamine release assay, and passive sensitization histamine release assay in the diagnosis of peanut allergy. *Allergy* 2018; 73:137-44.
38. Iweala OI, Choudhary SK, Addison CT, Batty CJ, Kapita CM, Amelio C, et al. Glycolipid-mediated basophil activation in alpha-gal allergy. *J Allergy Clin Immunol*. 2020 Aug;146(2):450-452. doi: 10.1016/j.jaci.2020.02.006. Epub 2020 Feb 20.
39. Moneret-Vautrin DA, Sainte-Laudy J, Kanny G, Fremont S. Human basophil activation measured by CD63 expression and LTC4 release in IgE-mediated food allergy. *Ann Allergy Asthma Immunol* 1999; 82:33-40.
40. Mueller-Wirth N, Buenter A, Jörg L, Ebo D, Glatz M, Fernando S, et al. IgE mediated Chlorhexidine Allergy - cross-reactivity with other biguanide disinfectants. *Allergy*; 2020 Jul 17. doi: 10.1111/all.14497. Online ahead of print.PMID: 32678912
41. Santos AF, Couto-Francisco N, Becares N, Kwok M, Bahnson HT, Lack G. A novel human mast cell activation test for peanut allergy. *J Allergy Clin Immunol* 2018; 142:689-91.e9.
42. Bahri R, Custovic A, Korosec P, Tsoumani M, Barron M, Wu J, et al. Mast cell activation test in the diagnosis of allergic disease and anaphylaxis. *J Allergy Clin Immunol* 2018; 142:485-96.e16.
43. Elst J, van der Poorten M-L, Faber M, Van Gasse A, Garvey L, Bridts C, et al. Mast cell activation tests: a proof-of-concept for a new asset in the mechanistic exploration and diagnosis of chlorhexidine immediate drug hypersensitivity. *Br J Anaesth*. 2020 Jul 21:S0007-0912(20)30482-7. doi: 10.1016/j.bja.2020.06.024. Online ahead of print.PMID: 32709306. For Editorial: Mast cell activation tests: a new tool in the investigation of suspected perioperative allergic reactions? Mertes PM, Hopkins PM.*Br J Anaesth*. 2020 Sep 25:S0007-0912(20)30726-1. doi: 10.1016/j.bja.2020.08.044. Online ahead of print.PMID: 32988603.
44. Bahri R, Bulfone-Paus S. Mast Cell Activation Test (MAT). *Methods Mol Biol* 2020; 2163:227-38.
45. Alenius H, Mäkinen-Kiljunen S, Ahlroth M, Turjanmaa K, Reunala T, Palosuo T. Crossreactivity between allergens in natural rubber latex and banana studied by immunoblot inhibition. *Clin Exp Allergy* 1996; 26:341-8.
46. Fahlbusch B, Rudeschko O, Muller WD, Schlenvoigt G, Vettermann S, Jäger L. Purification and characterization of the major allergen from apple and its allergenic cross-reactivity with Bet v 1. *Int Arch Allergy Immunol* 1995; 108:119-26.
47. Tomassen MM, Barrett DM, van der Valk HC, Woltering EJ. Isolation and characterization of a tomato non-specific lipid transfer protein involved in polygalacturonase-mediated pectin degradation. *J Exp Bot* 2007; 58:1151-60.
48. Larsen JN, Dreborg S. Standardization of Allergen Extracts. *Methods Mol Biol* 2019; 2020:63-76.
49. Sabato V, van Hengel AJ, De Knop KJ, Verweij MM, Hagendorens MM, Bridts CH, et al. Basophil activation reveals divergent patient-specific responses to thermally processed peanuts. *J Investig Allergol Clin Immunol* 2011; 21:527-31.
50. van der Veen MJ, van Ree R, Aalberse RC, Akkerdaas J, Koppelman SJ, Jansen HM, et al. Poor biologic activity of cross-reactive IgE directed to carbohydrate determinants of glycoproteins. *J Allergy Clin Immunol* 1997; 100:327-34.
51. Ebo DG, Lechkar B, Schuerwegh AJ, Bridts CH, De Clerck LS, Stevens WJ. Validation of a two-color flow cytometric assay detecting in vitro basophil activation for the diagnosis of IgE-mediated natural rubber latex allergy. *Allergy* 2002; 57:706-12.

- 557 52. Hoffmann HJ, Santos AF, Mayorga C, Nopp A, Eberlein B, Ferrer M, et al. The clinical utility of
558 basophil activation testing in diagnosis and monitoring of allergic disease. *Allergy* 2015;
559 70:1393-405.
- 560 53. Nopp A, Johansson SG, Ankerst J, Bylin G, Cardell LO, Gronneberg R, et al. Basophil allergen
561 threshold sensitivity: a useful approach to anti-IgE treatment efficacy evaluation. *Allergy*
562 2006; 61:298-302.
- 563 54. Santos AF, Du Toit G, Douiri A, Radulovic S, Stephens A, Turcanu V, et al. Distinct parameters
564 of the basophil activation test reflect the severity and threshold of allergic reactions to
565 peanut. *J Allergy Clin Immunol* 2015; 135:179-86.
- 566 55. Glaumann S, Nopp A, Johansson SG, Rudengren M, Borres MP, Nilsson C. Basophil allergen
567 threshold sensitivity, CD-sens, IgE-sensitization and DBPCFC in peanut-sensitized children.
568 *Allergy* 2012; 67:242-7.
- 569 56. Patil SU, Steinbrecher J, Calatroni A, Smith N, Ma A, Ruiter B, et al. Early decrease in basophil
570 sensitivity to Ara h 2 precedes sustained unresponsiveness after peanut oral immunotherapy.
571 *J Allergy Clin Immunol* 2019; 144:1310-9.e4.
- 572 57. Mayorga C, Andreu I, Aranda A, Dona I, Montanez MI, Blanca-Lopez N, et al. Fluoroquinolone
573 photodegradation influences specific basophil activation. *Int Arch Allergy Immunol* 2013;
574 160:377-82.
- 575 58. Watson BM, Oliveria JP, Nusca GM, Smith SG, Beaudin S, Dua B, et al. Inhibition of allergen-
576 induced basophil activation by ASM-024, a nicotinic receptor ligand. *Int Arch Allergy Immunol*
577 2014; 165:255-64.
- 578 59. Van Gasse AL, Elst J, Bridts CH, Mertens C, Faber M, Hagendorens MM, et al. Rocuronium
579 Hypersensitivity: Does Off-Target Occupation of the MRGPRX2 Receptor Play a Role? *J Allergy*
580 *Clin Immunol Pract* 2019; 7:998-1003.
- 581 60. Kepley CL, Youssef L, Andrews RP, Wilson BS, Oliver JM. Multiple defects in Fc epsilon RI
582 signaling in Syk-deficient nonreleaser basophils and IL-3-induced recovery of Syk expression
583 and secretion. *J Immunol* 2000; 165:5913-20.
- 584 61. Dreborg S. Methodological cutoff of basophil activation test and basophil activation test
585 diagnostic value. *J Allergy Clin Immunol Pract* 2018; 6:1089-90.
- 586 62. Gomez E, Campo P, Rondon C, Barrionuevo E, Blanca-Lopez N, Torres MJ, et al. Role of the
587 basophil activation test in the diagnosis of local allergic rhinitis. *J Allergy Clin Immunol* 2013;
588 132:975-6.e1-5.
- 589 63. Dahlen B, Nopp A, Johansson SG, Eduards M, Skedinger M, Adedoyin J. Basophil allergen
590 threshold sensitivity, CD-sens, is a measure of allergen sensitivity in asthma. *Clin Exp Allergy*
591 2011; 41:1091-7.
- 592 64. Lalek N, Kosnik M, Silar M, Korosec P. Immunoglobulin G-dependent changes in basophil
593 allergen threshold sensitivity during birch pollen immunotherapy. *Clin Exp Allergy* 2010;
594 40:1186-93.
- 595 65. Schmid JM, Wurtzen PA, Dahl R, Hoffmann HJ. Early improvement in basophil sensitivity
596 predicts symptom relief with grass pollen immunotherapy. *J Allergy Clin Immunol* 2014;
597 134:741-4.e5.
- 598 66. Sharif H, Singh I, Kouser L, Mosges R, Bonny MA, Karamani A, et al. Immunologic mechanisms
599 of a short-course of *Lolium perenne* peptide immunotherapy: A randomized, double-blind,
600 placebo-controlled trial. *J Allergy Clin Immunol* 2019; 144:738-49.
- 601 67. Kim SH, Kim SH, Chung SJ, Kim JH, Lee SY, Kim BK, et al. Changes in basophil activation during
602 immunotherapy with house dust mite and mugwort in patients with allergic rhinitis. *Asia Pac*
603 *Allergy* 2018; 8:e6.
- 604 68. Feng M, Zeng X, Su Q, Shi X, Xian M, Qin R, et al. Allergen Immunotherapy-Induced
605 Immunoglobulin G4 Reduces Basophil Activation in House Dust Mite-Allergic Asthma
606 Patients. *Front Cell Dev Biol* 2020; 8:30.
- 607 69. Faber M, Sabato V, De Witte L, Van Gasse A, Hagendorens MM, Leysen J, et al. State of the
608 art and perspectives in food allergy (part I): diagnosis. *Curr Pharm Des* 2014; 20:954-63.

70. Santos AF, Lack G. Basophil activation test: food challenge in a test tube or specialist research tool? *Clin Transl Allergy* 2016; 6:10.
71. Commins SP. Diagnosis & management of alpha-gal syndrome: lessons from 2,500 patients. *Expert Rev Clin Immunol*. 2020 Jul;16(7):667-677. doi: 10.1080/1744666X.2020.1782745. Epub 2020 Jul 8.PMID: 32571129
72. Ocmant A, Mulier S, Hanssens L, Goldman M, Casimir G, Mascart F, et al. Basophil activation tests for the diagnosis of food allergy in children. *Clin Exp Allergy* 2009; 39:1234-45.
73. Decuyper, Il, Pascal M, Van Gasse AL, Mertens C, Diaz-Perales A, Araujo G, et al. Performance of basophil activation test and specific IgG4 as diagnostic tools in nonspecific lipid transfer protein allergy: Antwerp-Barcelona comparison. *Allergy*; 2020 Mar;75(3):616-624. doi: 10.1111/all.14040. Epub 2019 Oct 7.PMID: 31512256
74. Ebo DG, Hagendorens MM, Bridts CH, Schuerwegh AJ, De Clerck LS, Stevens WJ. Flow cytometric analysis of in vitro activated basophils, specific IgE and skin tests in the diagnosis of pollen-associated food allergy. *Cytometry B Clin Cytom* 2005; 64:28-33.
75. Santos AF, Douiri A, Becares N, Wu SY, Stephens A, Radulovic S, et al. Basophil activation test discriminates between allergy and tolerance in peanut-sensitized children. *J Allergy Clin Immunol* 2014; 134:645-52.
76. Mehlich J, Fischer J, Hilger C, Swiontek K, Morisset M, Codreanu-Morel F, et al. The basophil activation test differentiates between patients with alpha-gal syndrome and asymptomatic alpha-gal sensitization. *J Allergy Clin Immunol* 2019; 143:182-9.
77. Ruinemans-Koerts J, Schmidt-Hieltjes Y, Jansen A, Savelkoul HFJ, Plaisier A, van Setten P. The Basophil Activation Test reduces the need for a food challenge test in children suspected of IgE-mediated cow's milk allergy. *Clin Exp Allergy* 2019; 49:350-6.
78. Erdmann SM, Sachs B, Schmidt A, Merk HF, Scheiner O, Moll-Slowsky S, et al. In vitro analysis of birch-pollen-associated food allergy by use of recombinant allergens in the basophil activation test. *Int Arch Allergy Immunol* 2005; 136:230-8.
79. Sato S, Tachimoto H, Shukuya A, Kurosaka N, Yanagida N, Utsunomiya T, et al. Basophil activation marker CD203c is useful in the diagnosis of hen's egg and cow's milk allergies in children. *Int Arch Allergy Immunol* 2010; 152 Suppl 1:54-61.
80. Chinthrajah RS, Purington N, Andorf S, Long A, O'Laughlin KL, Lyu SC, et al. Sustained outcomes in oral immunotherapy for peanut allergy (POISED study): a large, randomised, double-blind, placebo-controlled, phase 2 study. *Lancet* 2019; 394:1437-49.
81. Kim EH, Yang L, Ye P, Guo R, Li Q, Kulis MD, et al. Long-term sublingual immunotherapy for peanut allergy in children: Clinical and immunologic evidence of desensitization. *J Allergy Clin Immunol* 2019; 144:1320-6.e1.
82. Tsai M, Mukai K, Chinthrajah RS, Nadeau KC, Galli SJ. Sustained successful peanut oral immunotherapy associated with low basophil activation and peanut-specific IgE. *J Allergy Clin Immunol* 2020; 145:885-96.e6.
83. Nachshon L, Goldberg MR, Levy MB, Appel MY, Epstein-Rigbi N, Lidholm J, et al. Efficacy and Safety of Sesame Oral Immunotherapy-A Real-World, Single-Center Study. *J Allergy Clin Immunol Pract* 2019; 7:2775-81.e2.
84. Goldberg MR, Nachshon L, Appel MY, Elizur A, Levy MB, Eisenberg E, et al. Efficacy of baked milk oral immunotherapy in baked milk-reactive allergic patients. *J Allergy Clin Immunol* 2015; 136:1601-6.
85. Keet CA, Frischmeyer-Guerrerio PA, Thyagarajan A, Schroeder JT, Hamilton RG, Boden S, et al. The safety and efficacy of sublingual and oral immunotherapy for milk allergy. *J Allergy Clin Immunol* 2012; 129:448-55, 55.e1-5.
86. Matsui T, Naito M, Tagami K, Tajima I, Teshigawara M, Makino A, et al. Changes in passively-sensitized basophil activation to alphaS1-casein after oral immunotherapy. *Immun Inflamm Dis* 2020.

87. Giavi S, Vissers YM, Muraro A, Lauener R, Konstantinopoulos AP, Mercenier A, et al. Oral immunotherapy with low allergenic hydrolysed egg in egg allergic children. *Allergy* 2016; 71:1575-84.
88. Acosta GS, Kinaciyan T, Kitzmuller C, Mobs C, Pfutzner W, Bohle B. IgE-blocking antibodies following SLIT with recombinant Mal d 1 accord with improved apple allergy. *J Allergy Clin Immunol* 2020 Oct;146(4):894-900.e2. doi: 10.1016/j.jaci.2020.03.015. Epub 2020 Apr 4.
89. Wanich N, Nowak-Wegrzyn A, Sampson HA, Shreffler WG. Allergen-specific basophil suppression associated with clinical tolerance in patients with milk allergy. *J Allergy Clin Immunol* 2009; 123:789-94.e20.
90. Deng S, Yin J. Clinical utility of basophil activation test in diagnosis and predicting severity of mugwort pollen-related peach allergy. *World Allergy Organ J* 2019; 12:100043.
91. Santos AF, Du Toit G, O'Rourke C, Becares N, Couto-Francisco N, Radulovic S, et al. Biomarkers of severity and threshold of allergic reactions during oral peanut challenges. *J Allergy Clin Immunol*. 2020 Aug;146(2):344-355. doi: 10.1016/j.jaci.2020.03.035. Epub 2020 Apr 18.
92. Chinthrajah RS, Purington N, Andorf S, Rosa JS, Mukai K, Hamilton R, et al. Development of a tool predicting severity of allergic reaction during peanut challenge. *Ann Allergy Asthma Immunol* 2018; 121:69-76.e2.
93. Song Y, Wang J, Leung N, Wang LX, Lisann L, Sicherer SH, et al. Correlations between basophil activation, allergen-specific IgE with outcome and severity of oral food challenges. *Ann Allergy Asthma Immunol* 2015; 114:319-26.
94. Reier-Nilsen T, Michelsen MM, Lodrup Carlsen KC, Carlsen KH, Mowinckel P, Nygaard UC, et al. Predicting reactivity threshold in children with anaphylaxis to peanut. *Clin Exp Allergy* 2018; 48:415-23.
95. Rubio A, Vivinus-Nebot M, Bourrier T, Saggio B, Albertini M, Bernard A. Benefit of the basophil activation test in deciding when to reintroduce cow's milk in allergic children. *Allergy* 2011; 66:92-100.
96. Ford LS, Bloom KA, Nowak-Wegrzyn AH, Shreffler WG, Masilamani M, Sampson HA. Basophil reactivity, wheal size, and immunoglobulin levels distinguish degrees of cow's milk tolerance. *J Allergy Clin Immunol* 2013; 131:180-6.e1-3.
97. Chapuis A, Thevenot J, Coutant F, Messaoudi K, Michaud E, Pereira B, et al. Ara h 2 basophil activation test does not predict clinical reactivity to peanut. *J Allergy Clin Immunol Pract* 2018; 6:1772-4.e1.
98. Hayen SM, Ehlers AM, den Hartog Jager CF, Garssen J, Knol EF, Knulst AC, et al. 2S protein Ara h 7.0201 has unique epitopes compared to other Ara h 7 isoforms and is comparable to 2S proteins Ara h 2 and 6 in basophil degranulation capacity. *Clin Exp Allergy* 2018; 48:890-7.
99. Gernez Y, Tirouvanziam R, Yu G, Ghosn EE, Reshamwala N, Nguyen T, et al. Basophil CD203c levels are increased at baseline and can be used to monitor omalizumab treatment in subjects with nut allergy. *Int Arch Allergy Immunol* 2011; 154:318-27.
100. Ebo DG, Hagendorens MM, Bridts CH, De Clerck LS, Stevens WJ. Hymenoptera venom allergy: taking the sting out of difficult cases. *J Investig Allergol Clin Immunol* 2007; 17:357-60.
101. Korosec P, Silar M, Erzen R, Celesnik N, Bajrovic N, Zidarn M, et al. Clinical routine utility of basophil activation testing for diagnosis of hymenoptera-allergic patients with emphasis on individuals with negative venom-specific IgE antibodies. *Int Arch Allergy Immunol* 2013; 161:363-8.
102. Eberlein B, Krischan L, Darsow U, Ollert M, Ring J. Double positivity to bee and wasp venom: improved diagnostic procedure by recombinant allergen-based IgE testing and basophil activation test including data about cross-reactive carbohydrate determinants. *J Allergy Clin Immunol* 2012; 130:155-61.
103. Sturm GJ, Jin C, Kranzelbinder B, Hemmer W, Sturm EM, Griesbacher A, et al. Inconsistent results of diagnostic tools hamper the differentiation between bee and vespid venom allergy. *PLoS One* 2011; 6:e20842.

104. Ebo DG, Hagendorens MM, Schuerwegh AJ, Beirens LM, Bridts CH, De Clerck LS, et al. Flow-assisted quantification of in vitro activated basophils in the diagnosis of wasp venom allergy and follow-up of wasp venom immunotherapy. *Cytometry B Clin Cytom* 2007; 72:196-203.
105. Ott H, Tenbrock K, Baron J, Merk H, Lehmann S. Basophil activation test for the diagnosis of hymenoptera venom allergy in childhood: a pilot study. *Klin Padiatr* 2011; 223:27-32.
106. Erdmann SM, Sachs B, Kwiecien R, Moll-Sladowy S, Sauer I, Merk HF. The basophil activation test in wasp venom allergy: sensitivity, specificity and monitoring specific immunotherapy. *Allergy* 2004; 59:1102-9.
107. Zitnik SE, Vesel T, Avcin T, Silar M, Kosnik M, Korosec P. Monitoring honeybee venom immunotherapy in children with the basophil activation test. *Pediatr Allergy Immunol* 2012; 23:166-72.
108. Celesnik N, Vesel T, Rijavec M, Silar M, Erzen R, Kosnik M, et al. Short-term venom immunotherapy induces desensitization of FcepsilonRI-mediated basophil response. *Allergy* 2012; 67:1594-600.
109. Nullens S, Sabato V, Faber M, Leysen J, Bridts CH, De Clerck LS, et al. Basophilic histamine content and release during venom immunotherapy: insights by flow cytometry. *Cytometry B Clin Cytom* 2013; 84:173-8.
110. Bidad K, Nawijn MC, van Oosterhout AJ, van der Heide S, Elberink JN. Basophil activation test in the diagnosis and monitoring of mastocytosis patients with wasp venom allergy on immunotherapy. *Cytometry B Clin Cytom* 2014; 86:183-90.
111. Kosnik M, Silar M, Bajrovic N, Music E, Korosec P. High sensitivity of basophils predicts side-effects in venom immunotherapy. *Allergy* 2005; 60:1401-6.
112. Eberlein-Konig B, Schmidt-Leidescher C, Behrendt H, Ring J. Predicting side-effects in venom immunotherapy by basophil activation? *Allergy* 2006; 61:897.
113. Bonadonna P, Zanotti R, Melioli G, Antonini F, Romano I, Lenzi L, et al. The role of basophil activation test in special populations with mastocytosis and reactions to hymenoptera sting. *Allergy* 2012; 67:962-5.
114. Rietveld MJ, Schreurs MW, Gerth van Wijk R, van Daele PL, Hermans MA. The Basophil Activation Test Is Not a Useful Screening Tool for Hymenoptera Venom-Related Anaphylaxis in Patients with Systemic Mastocytosis. *Int Arch Allergy Immunol* 2016; 169:125-9.
115. Ebo DG, Hagendorens MM, Bridts CH, De Clerck LS, Stevens WJ. Sensitization to cross-reactive carbohydrate determinants and the ubiquitous protein profilin: mimickers of allergy. *Clin Exp Allergy* 2004; 34:137-44.
116. Faber MA, Sabato V, Bridts CH, Nayak A, Beezhoud DH, Ebo DG. Clinical relevance of the Hevea brasiliensis lipid transfer protein Hev b 12. *J Allergy Clin Immunol* 2015; 135:1645-8.
117. Ebo DG, Swerts S, Sabato V, Hagendorens MM, Bridts CH, Jorens PG, et al. New food allergies in a European non-Mediterranean region: is Cannabis sativa to blame? *Int Arch Allergy Immunol* 2013; 161:220-8.
118. Decuyper, II, Van Gasse AL, Faber MA, Elst J, Mertens C, Rihs HP, et al. Exploring the Diagnosis and Profile of Cannabis Allergy. *J Allergy Clin Immunol Pract* 2019; 7:983-9.e5.
119. Decuyper, II, Faber MA, Lapeere H, Mertens C, Rihs HP, Van Gasse AL, et al. Cannabis allergy: A diagnostic challenge. *Allergy* 2018; 73:1911-4.
120. Decuyper, II, Rihs HP, Mertens CH, Van Gasse AL, Elst J, De Puyseleir L, et al. A new cannabis allergen in North-western Europe; the oxygen-evolving enhancer protein 2 (OEEP2). *J Allergy Clin Immunol Pract* 2020.
121. Fernandez TD, Ariza A, Palomares F, Montanez MI, Salas M, Martin-Serrano A, et al. Hypersensitivity to fluoroquinolones: The expression of basophil activation markers depends on the clinical entity and the culprit fluoroquinolone. *Medicine (Baltimore)* 2016; 95:e3679.
122. Fernandez TD, Torres MJ, Blanca-Lopez N, Rodriguez-Bada JL, Gomez E, Canto G, et al. Negativization rates of IgE radioimmunoassay and basophil activation test in immediate reactions to penicillins. *Allergy* 2009; 64:242-8.

123. Ebo DG, Elst J, Van Gasse AL, De Puyseleir L, Faber MA, Mayorga C, et al. Basophil Activation Experiments in Immediate Drug Hypersensitivity: More Than a Diagnostic Aid. *Methods Mol Biol.* 2020;2163:197-211. doi: 10.1007/978-1-0716-0696-4_16.PMID: 32766977.
124. Mayorga C, Ebo DG, Lang DM, Pichler WJ, Sabato V, Park MA, et al. Controversies in drug allergy: In vitro testing. *J Allergy Clin Immunol* 2019; 143:56-65.
125. Takazawa T, Sabato V, Ebo DG. In vitro diagnostic tests for perioperative hypersensitivity, a narrative review: potential, limitations, and perspectives. *Br J Anaesth* 2019; 123:e117-e25.
126. Campos L, Galvao VR, Kalil J, Castells M, Giavina-Bianchi P. BAT in the Diagnosis of Drug Allergy: a Novel Tool in Clinical Daily Practice? *Curr Allergy Asthma Rep* 2019; 19:20.
127. Leysen J, De Witte L, Sabato V, Faber M, Hagendorens M, Bridts C, et al. IgE-mediated allergy to pholcodine and cross-reactivity to neuromuscular blocking agents: Lessons from flow cytometry. *Cytometry B Clin Cytom* 2013; 84:65-70.
128. Aranda A, Mayorga C, Ariza A, Dona I, Rosado A, Blanca-Lopez N, et al. In vitro evaluation of IgE-mediated hypersensitivity reactions to quinolones. *Allergy* 2011; 66:247-54.
129. Van Gasse AL, Sabato V, Uytendaele AP, Elst J, Faber MA, Hagendorens MM, et al. Immediate moxifloxacin hypersensitivity: Is there more than currently meets the eye? *Allergy* 2017; 72:2039-43.
130. Ebo DG, Baldo BA, Van Gasse AL, Mertens C, Elst J, Sermeus L, et al. Anaphylaxis to sugammadex-rocuronium inclusion complex: An IgE-mediated reaction due to allergenic changes at the sugammadex primary rim. *J Allergy Clin Immunol Pract.* 2020 Apr;8(4):1410-1415.e3. doi: 10.1016/j.jaip.2019.11.018. Epub 2019 Nov 27.
131. Ebo DG, Bridts CH, Hagendorens MM, Mertens CH, De Clerck LS, Stevens WJ. Flow-assisted diagnostic management of anaphylaxis from rocuronium bromide. *Allergy* 2006; 61:935-9.
132. Uytendaele AP, Sabato V, Leysen J, Bridts CH, De Clerck LS, Ebo DG. Flowcytometric diagnosis of atracurium-induced anaphylaxis. *Allergy* 2014; 69:1324-32.
133. Giavina-Bianchi P, Galvao VR, Picard M, Caiado J, Castells MC. Basophil Activation Test is a Relevant Biomarker of the Outcome of Rapid Desensitization in Platinum Compounds-Allergy. *J Allergy Clin Immunol Pract* 2017; 5:728-36.
134. Thevenot J, Ferrier le Bouedec MC, Buisson A, Bommelaer G, D'Incan M, Rouzair P. Rapid Desensitization to Adalimumab Is Associated With Decreased Basophil Sensitivity. *J Invest Allergol Clin Immunol* 2019; 29:141-3.
135. Van Gasse AL, Hagendorens MM, Sabato V, Bridts CH, De Clerck LS, Ebo DG. IgE to Poppy Seed and Morphine Are Not Useful Tools to Diagnose Opiate Allergy. *J Allergy Clin Immunol Pract* 2015; 3:396-9.
136. Harrer A, Lang R, Grims R, Braitsch M, Hawranek T, Aberer W, et al. Diclofenac hypersensitivity: antibody responses to the parent drug and relevant metabolites. *PLoS One* 2010; 5:e13707.
137. Sabato V, Van Gasse A, Cop N, Claesen K, Decuyper I, Faber M, et al. The Mas-Related G Protein-Coupled Receptor MRGPRX2 Is Expressed on Human Basophils and up-Regulated upon Activation. *J Allergy Clin Immunol* 2017; 139:AB168.
138. Huston DP, Sabato V. Decoding the Enigma of Urticaria and Angioedema. *J Allergy Clin Immunol Pract* 2018; 6:1171-5.
139. Maurer M, Eyerich K, Eyerich S, Ferrer M, Guterthum J, Hartmann K, et al. Urticaria: Collegium Internationale Allergologicum (CIA) Update 2020. *Int Arch Allergy Immunol* 2020;1-13.
140. Irinyi B, Gyimesi E, Garaczi E, Bata ZS, Kemeny L, Zeher M, et al. Extended diagnostic value of autologous serum skin test and basophil CD63 expression assay in chronic urticaria. *Br J Dermatol* 2013; 168:656-8.
141. Netchiporouk E, Moreau L, Rahme E, Maurer M, Lejtenyi D, Ben-Shoshan M. Positive CD63 Basophil Activation Tests Are Common in Children with Chronic Spontaneous Urticaria and Linked to High Disease Activity. *Int Arch Allergy Immunol* 2016; 171:81-8.

142. Chen Q, Zhai Z, Xu J, Chen W, Chen S, Zhong H, et al. Basophil CD63 expression in chronic spontaneous urticaria: correlation with allergic sensitization, serum autoreactivity and basophil reactivity. *J Eur Acad Dermatol Venereol* 2017; 31:463-8.
143. Patil SU, Calatroni A, Schneider M, Steinbrecher J, Smith N, Washburn C, et al. Data-driven programmatic approach to analysis of basophil activation tests. *Cytometry B Clin Cytom* 2018; 94:667-73.

Legends of figures main text

Figure 1: basophil identification and activation/degranulation markers

Basophils express various molecules that can be used in isolation or in combination to identify the cells and measure their activation/degranulation status by flow cytometry in a technique designated as basophil activation test (BAT). Basophil activation/degranulation can occur via IgE/FcεI-dependent and IgE/FcεRI-independent pathways. Individual cell activation/degranulation can be measured flow cytometrically on 4 levels. First, via appearance or up-regulation of surface markers (such as CD203c, CD63, CD300a, MRGPRX2 and avidin binding of membrane-associated exteriorized anionic proteoglycans released from granules). Second, via phosphorylation of signalling molecules (such as p38MAPK and STAT5). Third, via changes in mediator content (such as decrease of histamine or increase of trapped cytokines or their mRNA). Fourth, via increased intracellular calcium staining. Intracellular molecules are denoted in pink. For more details on the identification and activation/degranulation markers see text and Table E1 of repository. For technical details see ⁶⁻⁸.

Figure 2: Confocal microscopy image.

Confocal microscopy image of anti-IgE-stimulated basophils stained with anti-CD203c-APC, anti-surface IgE-AF405, anti-lysosomal associated protein CD63-PE, avidin-AF488 (binding membrane-associated exteriorized anionic proteoglycans released from granules) and composite view.

Figure 3: Allergen dose-response curve and TG-ROC analyses.

Dose-response (DR) curve for *ex vivo* basophil degranulation by recombinant Can s 3 (rCan s 3), the non-specific lipid transfer protein from *Cannabis sativa*, in 9 patients sensitized to Can s 3 (closed symbols). In contrast, in 7 control individuals cell are unresponsive to Can s 3 (open symbols). Results are obtained from individuals responsive to positive control stimulation and expressed as % CD63^{+ve} basophils, the dashed rectangle denotes 1 µg/mL to be the most discriminative stimulation concentration (left). Two-graph receiver operating characteristics curve combining sensitivity (squares) and specificity (circles) for the BAT rCan s 3 (1 µg/mL) (right). For a comparative validation of BAT rCan s 3 in cannabis allergy not restricted to Can s 3-sensitized patients see ¹¹⁸.

Figure 4: Resolving MRGPRX2- and IgE-binding properties of drugs: a guidance.

(a) Referring physicians should provide complete information and tryptase values. (b) Allergological work-up of immediate drug hypersensitivity generally starts with skin tests (STs). (c) Negative STs do not preclude sensitization. (d) Diagnosis of drug allergy rarely can rely upon a positive sIgE result in isolation. (e) ST negative cases have been reported in IgE-mediated drug allergy. An IgE-dependent reaction might be confirmed in a passive mast cell activation test (*indirect MAT). (f) In patients with evocative history and positive ST but incongruent negative combined sIgE, BAT and indirect MAT results may indicate an MRGPRX2-dependent reaction. (g) In cases with negative investigations drug challenges might be indicated. Note other mechanisms of IgE/FcεRI-independent MC activation by drugs can occur.

Legends of figures repository text

Figure E1: BAT sesame vs. sesame oil.

Individual plots for sesame and sesame oil in a patient who experienced urticaria after eating a salad dressed with sesame oil and anaphylaxis with hypotension during an open sesame challenge because of repetitive negative sIgE and skin test responses to sesame. Note the response to sesame oil (up-regulation of CD63 up to 35% and of CD203c up to 83%). For sesame no activation of the cells was demonstrable. Sesame extract was obtained using the extraction procedure described in ¹ and starting from crushed seeds.

Figure E2: BAT pholcodine and cross-reactive structures

Representative plots of basophil activation test (BAT) in a patient who experienced an immediate hypersensitivity reaction to the opiate antitussive pholcodine and who was uneventfully challenged with the structurally closely related compounds codeine and morphine. In the context of the pholcodine-hypothesis the patient was also tested for the aminosteroid-derived curarizing neuromuscular blocking agent rocuronium. Skin tests and BAT rocuronium were negative. See ²⁻⁴. Basophils are selected as IgE^{+ve}/CD203c^{+ve} cells. Pholcodine induces a clear dose-dependent upregulation of the lysosomal degranulation marker CD63 (up to 52% of the cells). In contrast, upon stimulation of cells with codeine and morphine, expression of CD63 remains merely unchanged.

Figure E3: Expression of MRGPRX2 by mast cells and basophils

Intracellular and surface expression of the Mas-related G protein coupled receptor MRGPRX2 by CD117^{+ve}/CD203c^{+ve} mast cells (MCs) and IgE^{+ve}/CD203c^{+ve} basophils (A). Unlike MCs, resting basophils barely express MRGPRX2 on their surface (B). However, MRGPRX2 is quickly up-regulated by basophils in response to positive control stimulation with anti-IgE (C). For details on MRGPRX2 staining in basophils see ⁵.

Figure E4: Representative plots of auto-immune basophil activation test (AIBAT) for suspected auto-immune chronic spontaneous urticaria. Donor basophils are identified as CD123^{+ve}/HLADR^{-ve} cells. After incubation with control serum there is no upregulation of CD63 (AIBAT: negative). After incubation of cells with serum of a patient, 94% of the basophils up-regulate expression of CD63.

Table E1: Basophil identification, activation and degranulation makers (actualized from ⁶)

	Marker	Other cell expression		References ⁽³⁾
Identification markers ⁽¹⁾	CD193 (CCR3)	MCs, Th2 lymphocytes	CD3 required for differentiation	7-20
	CD203c (ENPP3)	MCs, CD34 ⁺ progenitors	widely used identification marker	13, 15, 20-28
	CD123 (IL-3 receptor)	plasmacytoid DC	HLADR is frequently used to discriminate between HLADR ⁺ DCs	8, 12, 24, 25, 28-36
	IgE	monocytes, DCs	expression can be highly variable	13, 28
	CD294 (CRTH2)	MCs, eosinophils, Th2 lymphocytes	Differentiation from eosinophils via SSC or T lymphocytes by CD3	13, 21, 22, 24, 26, 37
Activation markers	CD63 ⁽²⁾ (LAMP3)	MCs, macrophages, platelets	Widely used degranulation marker	5, 7-16, 20-22, 25, 26, 28-43
	CD203c ⁽²⁾ (ENPP3)	See above		5, 14, 20-26, 28, 34, 35, 37-42, 44-46
	CD107a (LAMP1)/CD107b (LAMP2)	MCs, T lymphocytes, NK	CD63 compartment	21-23, 38
	CD11b	Macrophages, PMN, NK	CD203c compartment	47, 48
	CD13	myeloid cells	CD203c compartment	23, 38-41
	CD16	NK, neutrophils, monocytes, macrophages		20
	CD164	CD34 ⁺ progenitors	CD203c compartment	11, 23, 38, 41
	CD300a/CD200R	MCs, NK	Inhibitory receptors	49-51
	Phosphorylated p38 MAPK	Various cell types	phosphorylated members of the signalosome	52-54
	Phosphorylated STAT5	Various cell types	phosphorylated members of the signalosome	55
	Histamine	MCs	staining via DAO	21, 28, 56-59
	Exteriorized granule matrix	MCs	staining via avidin	60-64
	Trapping cytokines (IL-4, IL-13) or mRNA	Th 2 lymphocytes		65, 66

MCs: mast cells; DCs: dendritic cells; NK: natural killer cells;

CCR3: eotaxin receptor; ENPP3: ectonucleotide pyrophosphatase/phosphodiesterase family member 3; CD123: interleukin receptor α chain, CRTH2: prostaglandin D2 receptor; MRGPRX2: Mas-related G protein-coupled receptor X2; LAMP: lysosomal associated membrane protein; CD11b: part of complement receptor 3 (CR3), CD13: surface aminopeptidase-N; CD16: Fc receptor (Fc γ RIII); CD164: sialomucin core protein 24; CD300a: inhibitory receptor p 60 (IRp60); CD200R: OX2 receptor; MAPK: mitogen-activated protein kinase; STAT: signal transducer and activator of transcription.

Intracellular histamine content can be quantified flow cytometrically by using histaminase diamine oxidase (DAO) that is conjugated to a fluorochrome.

Exteriorized proteoglycans can be stained by fluorescent avidin probes.

⁽¹⁾ Identification markers are generally used in combinations (see text).

⁽²⁾ CD203c compartment: rapid and significant upregulation of CD13, CD164 and CD203c; CD63 compartment: slower upregulation of CD107a and CD63 ³⁸.

⁽³⁾ Not exhaustive.

Santos et al, demonstrated that expression of CD123 can decrease upon activation of cells²⁵. However, the authors and others do not share this experience^{36,67}. Corvan et al, provide evidence for IgE-independent activation of CD203c⁺ KU 812 basophil-like cells by the nematode *Trichostrongylus colubriformis* (a parasite of herbivorous mammals)²³.

Journal Pre-proof

Table E2: BAT in inhalant allergy (data for CD63-BAT, except if denoted in <i>italics</i>)					
Allergen	Reference test	Sensitivity	Specificity	Number	Reference
House dust mite (HDM) <i>Dactylis glomerata</i>	H + IgE and/or ST	56-78 73-100	91-100 100	20	⁶⁸
Cypress pollen	H + ST + PT	91	100	75	⁶⁹
HDM and <i>Lolium perenne</i>	H + ST + PT	93	98	166	⁷⁰
nFel d 1	H + IgE and/or ST	100 95	95	39	⁷¹
nPhl p 5 (45.1 µg/mL)	ST	<i>100</i>	<i>100</i>	40	⁷²
H: history, ST: skin test, sIgE: specific IgE, PT: provocation test. <i>Italics</i> : CD203c-BAT. Predictive values are summarized in table E9.					

Table E3: BAT in food allergy (data for CD63-BAT, except if denoted in <i>italics</i>)					
Allergen	Reference test	Sensitivity	Specificity	Number	Reference
Various foods	H, ST, OFC	58	97	64	⁷³
Birch-celery	H (OAS)	85	80	49	⁷⁴
Birch-carrot		80	95		
Birch-hazelnut		90	90		
Birch-apple	H (OAS)	100 88 [§]	100 75 [§]	61	⁷⁵
Birch-rMal d 1 (apple)	H (OAS)	75 [£]	68 [£]	55	⁷⁶
Birch-rApi g 1 (celery)		75 [£]	77 [£]		
Birch-rDau c 1 (carrot)		65 [£]	100 [£]		
<i>Macrobrachium Rosenbergii</i>	H	100	100	45	⁷⁷
Peanut	H + sIgE and/or ST	87 90	94 97	75	⁷⁸
Egg	H + sIgE and/or ST	89 63	100 97	68	⁷⁸
Wheat PBS fraction	OFC or severe proven H	86	58		⁴⁵
Wheat ethanol fraction		83	69		
Wheat alkaline fraction		84	67		
Native Ω-5 gliadine		85	77		
Recombinant Ω-5 gliadine		82	63		
Cow's milk	OFC or recent convincing H	89	83	Total of 71 (milk and egg)	⁴⁶
Casein		67	71		
Egg's white (HEA+ vs HEA-)		74	62		
Egg's white (REA+ vs REA-)		77	63		
Ovomucoid (HEA+ vs HEA-)		80	73		
Ovomucoid (REA+ vs REA-)		83	83		
Peanut	H and sIgE and ST	92	95	45	⁷⁹
Ara h 9		88	100		
Sesame	OFC or recent convincing H	86 84	85 80	82	⁸⁰
Peanut	DBPCFC	92	77	34	⁸¹
Peanut	OFC or severe proven H	98 95	96 96	104	⁸²
Cow's milk	DBPCFC	100	100		⁸³
Pru p 3 (Antwerp)	H and sIgE	NAV	NAV	219	⁸⁴
Pru p 3 (Barcelona)		NAV	NAV		
Mal d 3 (Antwerp)		NAV	NAV		
Mal d 3 (Barcelona)		63	67		
Alpha-gal syndrome	No sensitivity and specificity data available. However, BAT discriminates between asymptomatic sensitization and patients with a genuine alpha-gal syndrome.				¹⁶

H: history, ST: skin test, sIgE: specific IgE, OFC: oral food challenge, DBPCFC: double blind placebo controlled food challenge. HEA: heated egg allergic, REA: raw egg allergic. *Italics*: CD203c-BAT.

[£] Lower sensitivity in ⁷⁶ as compared to ⁷⁴ is likely to reflect the Bet v 1 homologues (Api g 1 and Dau c 1) not to cover the entire sensitisation profile. The same holds true for casein that does not cover the cow's milk reactivity profile but ovomucoid vs. egg's white ⁴⁶.

[§] according to a separate analysis between birch pollen allergic patients with and without apple-related oral allergy syndrome (OAS).

The study by Tokuda et al ⁴⁵ shows allergen-specific decision thresholds.

The study by Decuyper et al ⁸⁴ emphasizes the importance of geographically distinct sensitization profiles and component-related differences. NAV: no added value for diagnosis.
For additional information on BAT in food allergy see: ⁸⁵.
Predictive values are summarized in table E9.

Journal Pre-proof

Table E4: BAT in hymenoptera venom allergy (data for CD63-BAT, except if denoted in <i>italics</i>)					
Allergen	Reference test	Sensitivity	Specificity	Number	Reference
Wasp and honeybee	H	100	100	55	⁸⁶
Wasp and honeybee	H	<i>91</i>	<i>85</i>	35	⁸⁷
Wasp (Ves v 5)	H	<i>79</i>	<i>100</i>	34	⁸⁸
Honeybee (Api m 1)		<i>69</i>	<i>100</i>		
Wasp and honeybee	H and ST	77	90	44	⁸⁹
Wasp	H	85	87 (overall)	34	⁹⁰
Honeybee		91		23	
Wasp	H	92	80	70	⁹¹
Wasp	H + ST + sIgE	97	100	38	⁹²
Honeybee		100	100	12	
Wasp and honeybee	H + ST + sIgE	89	100	68	⁹³
		<i>97</i>	<i>89</i>		
Wasp	H	84	100	94	⁹⁴
Wasp	H	87	97	178	⁹⁵
Honeybee		90	95		
Wasp and honeybee	H	90		204	⁹⁶
Wasp and honeybee (children)	H	67-75		15	⁹⁷

H: history, ST: skin test, sIgE: specific IgE. *Italics*: CD203c-BAT

Wasp venom allergen 5 (Ves v 5) and honeybee venom phospholipase A2 (Api m 1) do not cover the entire IgE reactivity profile of wasp and honeybee venom, respectively ⁸⁸. Overall sensitivity of BAT using Ves v 1, Ves v 2, Ves v 5, Api m 1, Api m 2 and Api m 4 was 90%.

From the study by Eberlein-König et al ⁹², it appears BAT to be highly utile. However, it is poorly discriminative, as 75% of honeybee venom allergic patients had a positive BAT wasp, and, vice versa, in patients with wasp venom allergy 55% had a positive BAT for honeybee venom.

Note, as indicated in the manuscript, in patients with mastocytosis the BAT adds little to the diagnosis in difficult cases with negative sIgE and ST results ^{98,99}.

Predictive values are summarized in table E9.

Table E5: BAT in <i>Hevea latex</i> allergy (data for CD63-BAT, except if denoted in <i>italics</i>)					
Allergen	Reference test	Sensitivity	Specificity	Number	Reference
Latex	H + sIgE + ST	93	92	102*	¹⁰⁰
Latex	H + ST	93	100	73	¹⁰¹
Latex	H + ST	50 75	100 100	28	¹⁰²
Latex	H + ST + GPT	95	100	43	¹⁰³
Latex	H + sIgE + ST	80	97	79	¹⁰⁴
Latex rHev b 5, rHev b 6	H + sIgE + ST	95	100	33	¹⁰⁵
H: history, ST: skin test, sIgE: specific IgE, PT: glove provocation test. <i>Italics</i>: CD203c-BAT * Includes atopic control individuals. <i>Hevea brasiliensis</i> (Hev b) is the rubber tree. Predictive values are summarized in table E9.					

Table E6: BAT in immediate β -lactam hypersensitivity (data for CD63-BAT, except if denoted in *italics*). Actualized from ¹⁰⁶.

Stimulus	Reference test	Activation marker	Sensitivity	Specificity	Number of patients and controls	Ref.
β -lactams	H	CD63	50	93	88	¹⁰⁷
β -lactams	H + DPT ¹	CD63	39	93	53	¹⁰⁸
β -lactams	H $\pm$ ST $\pm$ slgE $\pm$ DPT	CD63	49	91	110	¹⁰⁹
Amoxicillin	H $\pm$ ST	CD203c	52	100	41	¹¹⁰
		CD63	22	79		
β -lactams	H	CD63	50	89-97	262	¹¹¹
β -lactams	H $\pm$ ST $\pm$ slgE	CD63-CCR3	55	100	39	¹⁰
		CD63-IgE	53			
Amoxicillin	H	CD63	29	/	14 patients, no controls	¹¹²
Amoxicillin	H $\pm$ ST $\pm$ DPT	CD63	50	/	61 patients, number of controls not mentioned	¹¹³
Amoxicillin	H $\pm$ ST	CD63	50	/	30 patients	¹¹⁴
Amoxicillin Clavulanic acid	H $\pm$ ST $\pm$ DPT	CD63	47	93	57	¹¹⁵
			62	89	58	
Cefazolin	H + ST	CD63	33	94	16 patients,	¹¹⁶
		<i>CD203c</i>	<i>67</i>	<i>94</i>	17 controls	

H: history, ST: skin test, DPT: drug provocation test, N: number of patients and control individuals. *Italics*: CD203c-BAT.

Note that several studies were not restricted to a single compound but investigated BAT in a drug class . For comment on ¹¹⁵ see ¹¹⁷.

Predictive values are summarized in table E9. See also ¹¹⁸.

Table E7: BAT in immediate neuromuscular blocking agent (NMBA) hypersensitivity (data for CD63-BAT, except if denoted in *italics*). Actualized from ¹⁰⁶.

Stimulus	Reference Test	Activation marker	Sensitivity	Specificity	N	Ref.
NMBAs	H + ST	CD63	64	81	26	¹¹⁹
		CD45	43	96		
NMBAs	H ± ST	CD63	54	100	56	⁷
NMBAs	H + ST	CD63	79	100	31	¹²⁰
		<i>CD203c</i>	<i>36</i>	<i>100</i>		
NMBAs	H ± ST	CD63	36-86 ¹	93	92	¹²¹
Rocuronium	H + ST	CD63	92 ²	100	22	³⁰
NMBAs	H ± ST ± slgE	CD63	60	100	92	¹²²
Rocuronium	H + ST	CD63	80	96	104	³¹
NMBAs	H+ST	CD63	68	100	56	¹²³
Atracurium	H + ST	CD63	71 ³	100	75	¹²⁴
Rocuronium	H + ST	<i>DAO</i>	100	100	13	⁵⁹
NMBAs ^{4,5}	H ± ST ± slgE	CD63	48	87	120	¹²⁵
		<i>CD203c</i>	<i>58</i>	<i>89</i>		

H: history, ST: skin test, N: number of patients and control individuals. *Italics*: CD203c-BAT

¹ Increasing sensitivity when only the reactions that occurred during the 3 years were taken into account,

² taking into account the non-responders sensitivity is 76%,

³ taking into account the non-responders sensitivity is 63%.

⁴ sensitivity and specificity for %CD63 and %CD203c. Overall sensitivity and specificity is 77% and 76%, respectively.

⁵ BAT also positive in some patients with negative skin tests (note that CD300a was used in some cases). BAT was also found to supplement skin tests in Van Gasse et al ¹²⁶.

DAO: conjugated diamine oxidase is used to measure intracellular histamine content (proof-of-concept in IDHR) ⁵⁹.

Predictive values are summarized in table E9. See also ¹¹⁸.

Table E8: BAT fluoroquinolone (FQ) hypersensitivity (data for CD63-BAT, except if denoted in *italics*). Actualized from ¹⁰⁶.

Stimulus	Reference Test	Activation marker	Sensitivity	Specificity	N	Ref.
FQs	H + DPT	CD63	0	–	4	¹²⁷
FQs	H + ST + DPT	CD63	0	100	18	¹²⁸
FQs	H	<i>CD203c</i>	<i>100</i>	<i>100</i>	5	¹²⁹
FQs	H + DPT	CD63	71	–	75	¹³⁰
FQs	H + DPT	<i>CD203c</i>	NA	<i>100</i>	34	¹³¹
Moxifloxacin	H + DPT	CD63	9.1	77.8	11	¹⁴
		<i>CD203c</i>	<i>36.4</i>	<i>94.4</i>	11	
Ciprofloxacin		CD63	83.8	88.9	6	
		<i>CD203c</i>	<i>0</i>	<i>94.4</i>	6	
Moxifloxacin	H	CD63	13.3	100	24	¹³²
		<i>CD203c</i>	<i>46.7</i>	<i>100</i>	24	

H: history, ST: skin test, DPT: drug provocation test, N: number of patients and control individuals
Italics: CD203c-BAT
 Predictive values are summarized in table E9. See also ¹¹⁸.

Table E9: positive and negative predictive values of BAT (not exhaustive, for drug also see ¹¹⁸)					
Allergen	Reference test	PPV (%)	NPV (%)	Numbers	References
Inhalant allergens					
HDM and Lolium perenne	H + ST + PT	99	89.7	166	⁷⁰
Food allergens					
Various wheat fractions	OFC or severe proven H	77.1-86.8	66.7-72	58	⁴⁵
Ovomucoid	OFC or recent convincing H	100	31.3	Total of 71 (milk and egg)	⁴⁶
Egg's white		94.7	21.9		
Cow's milk		85.7	86.4		
Casein		75	59.1		
Peanut	OFC or severe proven H	95	98	104	⁸²
Drugs					
β -lactams	H	93.5	49.1	88	¹⁰⁷
β -lactams	H $\pm$ ST $\pm$ slgE $\pm$ DPT	50.8	92.4	110	¹⁰⁹
Amoxicillin	H $\pm$ ST	85.2	52.5	41	¹¹⁰
β -lactams	H	90.5	43.8	262	¹¹¹
β -lactams	H $\pm$ ST $\pm$ slgE	81.5	52.6	39	¹⁰
FQs	H + DPT	90.1	66.7	75	¹³⁰
FQs	H + DPT	100	78.9	34	¹³¹
Moxifloxacin	H + DPT	81.8	66.7	35	¹⁴
Ciprofloxacin					
NMBAs	H + ST	86.9	78.1	26	¹¹⁹
NMBAs	H + ST	100	48.6	56	⁷
NMBAs	H + ST	100	76.9	31	¹²⁰
Rocuronium	H + ST	100	87.3	22	³⁰
NMBAs	H $\pm$ ST $\pm$ DPT	84.9	58.3	92	¹²¹
Rocuronium	H + ST + IgE	97	75	104	³¹
NMBAs	H + ST	100	82.9	56	¹²³
Atracurium	H + ST	100	70	75	¹²⁴
In the study by Sanz et al ⁷⁰, the likelihood ratio for a positive value was 58.3, for a negative result 0.07 In the study by Tokuda et al ⁴⁵, the positive predictive value (PPV) and negative predictive value (NPV) for slgE wheat by ImmunoCAP (Thermo Fisher Scientific) was 74.5% and 65.2%, respectively. In the study by Leysen et al ³¹, the PPV for ST and slgE rocuronium was 98% and 83%; the NPV for ST and slgE rocuronium was 96% and 72%.					

Table E10: Basophil activation tests: requirements and points to consider to advance further entrance in mainstream use
Technical
Standardization between systems and instruments (required for accreditation) - E.g. Euroflow standard operating procedure.
Standardization of sample preparations (extraction procedure, storage, preservatives) - E.g. aqueous vs. alcoholic extraction procedures, lyophilization, spiking with purified or recombinants - For recombinants: cave presence of maltose binding protein (MBP) See also: Larsen et al. Methods Mol Biol. 2019;2020:63-76. doi: 10.1007/978-1-4939-9591-2_5.
Sample conditions: chelators or heparin, temperature, preservation, whole blood vs. purified cells - E.g. when EDTA is used as anticoagulant correct Ca restoring is mandatory to avoid false negative outcomes
Stimulation conditions (dose-findings (with reporting of end concentrations in aliquot), incubation time, incubation temperature (prewarming of reagents), use of primers (e.g. IL-3), stopping by chelators - For drugs: take into account pharmacokinetics, cytotoxicity (viability should be monitored), protein binding, pharmacologic inhibition (e.g. via nicotine receptors), use of metabolites, photodegradation, interaction with fluorochromes, stimulation concentrations are best expressed in $\mu\text{mol/L}$
Calculation of allergen-specific decision thresholds (absolute % of positive cells, MFI, stimulation indices). Arbitrarily chosen cut-off should be abandoned. In BAT traditional cut-off, defined as result of blank (stimulation with buffer) + 3.3 SD, is rarely adopted.
Non-responders should be reported and included in performance analyses Basophils can show inherent biological variability from one day to another - Repetition of analyses could be worthwhile - Whether passive BAT using stripped donor basophils, humanized rat basophilic leukemic cell lines (e.g. RBL-2H3), tumoral MC lines (e.g. LAD2) or mast cell activation tests (MAT) using cultured donor MCs remains to be established
Automation and introduction of bioinformatics should improve efficiency, transparency and standardization.
Clinical
Identification and enrollment of sufficient numbers of well-documented cases and exposed control individuals. - Consensus about mechanistic nomenclature - Consensus scores about clinical presentation and severity - Application of gold standard or reference test (e.g. challenges)
Establishment of sensitization profile. - E.g. not allergens are equally potent
Identification of confounders - E.g. underlying cross-reactivity

References (of the repository file)

1. Alenius H, Mäkinen-Kiljunen S, Ahlroth M, Turjanmaa K, Reunala T, Palosuo T. Crossreactivity between allergens in natural rubber latex and banana studied by immunoblot inhibition. *Clin Exp Allergy* 1996; 26:341-8.
2. Leysen J, De Witte L, Sabato V, Faber M, Hagendorens M, Bridts C, et al. IgE-mediated allergy to pholcodine and cross-reactivity to neuromuscular blocking agents: Lessons from flow cytometry. *Cytometry B Clin Cytom* 2013; 84:65-70.
3. Swerts S, Van Gasse A, Leysen J, Faber M, Sabato V, Bridts CH, et al. Allergy to illicit drugs and narcotics. *Clin Exp Allergy* 2014; 44:307-18.
4. Van Gasse AL, Hagendorens MM, Sabato V, Bridts CH, De Clerck LS, Ebo DG. IgE to Poppy Seed and Morphine Are Not Useful Tools to Diagnose Opiate Allergy. *J Allergy Clin Immunol Pract* 2015; 3:396-9.
5. Elst J, Sabato V, Hagendorens MM, van Houdt M, Faber MA, Bridts CH, et al. Measurement and Functional Analysis of the Mas-Related G Protein-Coupled Receptor MRGPRX2 on Human Mast Cells and Basophils. *Methods Mol Biol* 2020; 2163:219-26.
6. Hemmings O, Kwok M, McKendry R, Santos AF. Basophil Activation Test: Old and New Applications in Allergy. *Curr Allergy Asthma Rep* 2018; 18:77.
7. Monneret G, Benoit Y, Debard AL, Gutowski MC, Topenot I, Bienvenu J. Monitoring of basophil activation using CD63 and CCR3 in allergy to muscle relaxant drugs. *Clin Immunol* 2002; 102:192-9.
8. Ducrest S, Meier F, Tschopp C, Pavlovic R, Dahinden CA. Flowcytometric analysis of basophil counts in human blood and inaccuracy of hematology analyzers. *Allergy* 2005; 60:1446-50.
9. Wolanczyk-Medrala A, Gogolewski G, Liebhart J, Gomulka K, Litwa M, Panaszek B, et al. A new variant of the basophil activation test for allergen-induced basophil CD63 upregulation. The effect of cetirizine. *J Investig Allergol Clin Immunol* 2009; 19:465-73.
10. Eberlein B, Leon Suarez I, Darsow U, Rueff F, Behrendt H, Ring J. A new basophil activation test using CD63 and CCR3 in allergy to antibiotics. *Clin Exp Allergy* 2010; 40:411-8.
11. Wolanczyk-Medrala A, Barg W, Liebhart J, Panaszek B, Nadobna G, Litwa M, et al. Validation of basophil CD164 upregulation for pollen allergy diagnosis. *Arch Immunol Ther Exp (Warsz)* 2010; 58:459-65.
12. Hausmann OV, Gentinetta T, Fux M, Ducrest S, Pichler WJ, Dahinden CA. Robust expression of CCR3 as a single basophil selection marker in flow cytometry. *Allergy* 2011; 66:85-91.
13. Eberlein B, Hann R, Eyerich S, Pennino D, Ring J, Schmidt-Weber CB, et al. Optimizing of the basophil activation test: Comparison of different basophil identification markers. *Cytometry B Clin Cytom* 2015; 88:183-9.
14. Fernandez TD, Ariza A, Palomares F, Montanez MI, Salas M, Martin-Serrano A, et al. Hypersensitivity to fluoroquinolones: The expression of basophil activation markers depends on the clinical entity and the culprit fluoroquinolone. *Medicine (Baltimore)* 2016; 95:e3679.
15. Depince-Berger AE, Sidi-Yahya K, Jeraiby M, Lambert C. Basophil activation test: Implementation and standardization between systems and between instruments. *Cytometry A* 2017; 91:261-9.
16. Mehlich J, Fischer J, Hilger C, Swiontek K, Morisset M, Codreanu-Morel F, et al. The basophil activation test differentiates between patients with alpha-gal syndrome and asymptomatic alpha-gal sensitization. *J Allergy Clin Immunol* 2019; 143:182-9.
17. Urrea JM, Perez-Lucendo I, Extremera A, Feo-Brito F, Alfaya T. The Method for Selecting Basophils Might Be Determinant in the Basophil Activation Test in Patients With Mastocytosis. *J Investig Allergol Clin Immunol* 2020; 30:65-7.
18. Chirumbolo S. Basophil Activation Test for Chronic Urticaria. *Ann Lab Med* 2016; 36:499-500.

19. Chirumbolo S, Bjorklund G, Vella A. Bias in the use of a SSC(low)/CCR3(pos) gate to capture basophils in chronic urticaria? *Immunobiology* 2019; 224:353-4.
20. Heneberg P, Riegerova K, Rihova A, Simcikova D, Kucera P. Updates on the surface antigens of basophils: CD16 on basophils of patients with respiratory or insect venom allergy and the rejection of CD203c and CD63 externalization decoupling by bisindolylmaleimides. *Clin Exp Allergy* 2019; 49:54-67.
21. Shamji MH, Layhadi JA, Scadding GW, Cheung DK, Calderon MA, Turka LA, et al. Basophil expression of diamine oxidase: a novel biomarker of allergen immunotherapy response. *J Allergy Clin Immunol* 2015; 135:913-21.e9.
22. Shamji MH, Bellido V, Scadding GW, Layhadi JA, Cheung DK, Calderon MA, et al. Effector cell signature in peripheral blood following nasal allergen challenge in grass pollen allergic individuals. *Allergy* 2015; 70:171-9.
23. Corvan SM, Agnew L, Andronicos NM. *Trichostrongylus colubriformis* induces IgE-independent CD13, CD164 and CD203c mediated activation of basophils in an in vitro intestinal epithelial cell co-culture model. *Vet Parasitol* 2015; 207:285-96.
24. Gernez Y, Tirouvanziam R, Yu G, Ghosn EE, Reshamwala N, Nguyen T, et al. Basophil CD203c levels are increased at baseline and can be used to monitor omalizumab treatment in subjects with nut allergy. *Int Arch Allergy Immunol* 2011; 154:318-27.
25. Santos AF, Becares N, Stephens A, Turcanu V, Lack G. The expression of CD123 can decrease with basophil activation: implications for the gating strategy of the basophil activation test. *Clin Transl Allergy* 2016; 6:11.
26. Sharif H, Singh I, Kouser L, Mosges R, Bonny MA, Karamani A, et al. Immunologic mechanisms of a short-course of *Lolium perenne* peptide immunotherapy: A randomized, double-blind, placebo-controlled trial. *J Allergy Clin Immunol* 2019; 144:738-49.
27. Nopp A, Johansson SG, Ankerst J, Bylin G, Cardell LO, Gronneberg R, et al. Basophil allergen threshold sensitivity: a useful approach to anti-IgE treatment efficacy evaluation. *Allergy* 2006; 61:298-302.
28. Bridts CH, Sabato V, Mertens CH, Hagendorens MM, De Clerck LS, Ebo DG. Flow Cytometric Allergy Diagnosis: Basophil Activation Techniques. *Methods Mol Biol.* 2020;2163:183-195. doi: 10.1007/978-1-0716-0696-4_15.PMID: 32766976.
29. Gonzalez-Munoz M, Luque R, Nauwelaers F, Moneo I. Detection of *Anisakis simplex*-induced basophil activation by flow cytometry. *Cytometry B Clin Cytom* 2005; 68:31-6.
30. Ebo DG, Bridts CH, Hagendorens MM, Mertens CH, De Clerck LS, Stevens WJ. Flow-assisted diagnostic management of anaphylaxis from rocuronium bromide. *Allergy* 2006; 61:935-9.
31. Leysen J, Bridts CH, De Clerck LS, Vercauteren M, Lambert J, Weyler JJ, et al. Allergy to rocuronium: from clinical suspicion to correct diagnosis. *Allergy* 2011; 66:1014-9.
32. Frezzolini A, Provini A, Teofoli P, Pomponi D, De Pita O. Serum-induced basophil CD63 expression by means of a tricolour flow cytometric method for the in vitro diagnosis of chronic urticaria. *Allergy* 2006; 61:1071-7.
33. Sturm GJ, Kranzelbinder B, Sturm EM, Heinemann A, Groselj-Strele A, Aberer W. The basophil activation test in the diagnosis of allergy: technical issues and critical factors. *Allergy* 2009; 64:1319-26.
34. Wanich N, Nowak-Wegrzyn A, Sampson HA, Shreffler WG. Allergen-specific basophil suppression associated with clinical tolerance in patients with milk allergy. *J Allergy Clin Immunol* 2009; 123:789-94.e20.
35. Ford LS, Bloom KA, Nowak-Wegrzyn AH, Shreffler WG, Masilamani M, Sampson HA. Basophil reactivity, wheal size, and immunoglobulin levels distinguish degrees of cow's milk tolerance. *J Allergy Clin Immunol* 2013; 131:180-6.e1-3.
36. Chirumbolo S, Bjorklund G, Vella A. Using a CD45dim/CD123bright/HLA-DRneg phenotyping protocol to gate basophils in FC for airway allergy. CD123 does not decrease. *Adv Respir Med* 2017; 85:193-201.

37. Horiuchi T, Yokohama A, Orihara M, Tomita Y, Tomioka A, Yoshida N, et al. Usefulness of Basophil Activation Tests for Diagnosis of Sugammadex-Induced Anaphylaxis. *Anesth Analg* 2018; 126:1509-16.
38. Hennersdorf F, Florian S, Jakob A, Baumgartner K, Sonneck K, Nordheim A, et al. Identification of CD13, CD107a, and CD164 as novel basophil-activation markers and dissection of two response patterns in time kinetics of IgE-dependent upregulation. *Cell Res* 2005; 15:325-35.
39. Chirumbolo S, Vella A, Ortolani R, De Gironcoli M, Solero P, Tridente G, et al. Differential response of human basophil activation markers: a multi-parameter flow cytometry approach. *Clin Mol Allergy* 2008; 6:12.
40. Sonneck K, Baumgartner C, Rebuzzi L, Marth K, Chen KW, Hauswirth AW, et al. Recombinant allergens promote expression of aminopeptidase-n (CD13) on basophils in allergic patients. *Int J Immunopathol Pharmacol* 2008; 21:11-21.
41. Kneidinger M, Schmidt U, Rix U, Gleixner KV, Vales A, Baumgartner C, et al. The effects of dasatinib on IgE receptor-dependent activation and histamine release in human basophils. *Blood* 2008; 111:3097-107.
42. Hauswirth AW, Sonneck K, Florian S, Krauth MT, Bohm A, Sperr WR, et al. Interleukin-3 promotes the expression of E-NPP3/CD203C on human blood basophils in healthy subjects and in patients with birch pollen allergy. *Int J Immunopathol Pharmacol* 2007; 20:267-78.
43. Ebo DG, Baldo BA, Van Gasse AL, Mertens C, Elst J, Sermeus L, et al. Anaphylaxis to sugammadex-rocuronium inclusion complex: An IgE-mediated reaction due to allergenic changes at the sugammadex primary rim. *J Allergy Clin Immunol Pract* 2019.
44. Smiljkovic D, Kiss R, Lupinek C, Hoermann G, Greiner G, Witzeneder N, et al. Microarray-based Detection of Allergen-Reactive IgE in Patients with Mastocytosis. *J Allergy Clin Immunol Pract* 2020 Sep;8(8):2761-2768.e16. doi: 10.1016/j.jaip.2020.04.030.
45. Tokuda R, Nagao M, Hiraguchi Y, Hosoki K, Matsuda T, Kouno K, et al. Antigen-induced expression of CD203c on basophils predicts IgE-mediated wheat allergy. *Allergol Int* 2009; 58:193-9.
46. Sato S, Tachimoto H, Shukuya A, Kurosaka N, Yanagida N, Utsunomiya T, et al. Basophil activation marker CD203c is useful in the diagnosis of hen's egg and cow's milk allergies in children. *Int Arch Allergy Immunol* 2010; 152 Suppl 1:54-61.
47. Yoshimura C, Yamaguchi M, Iikura M, Izumi S, Kudo K, Nagase H, et al. Activation markers of human basophils: CD69 expression is strongly and preferentially induced by IL-3. *J Allergy Clin Immunol* 2002; 109:817-23.
48. MacGlashan D, Jr. Marked differences in the signaling requirements for expression of CD203c and CD11b versus CD63 expression and histamine release in human basophils. *Int Arch Allergy Immunol* 2012; 159:243-52.
49. Sabato V, Verweij MM, Bridts CH, Levi-Schaffer F, Gibbs BF, De Clerck LS, et al. CD300a is expressed on human basophils and seems to inhibit IgE/FcepsilonRI-dependent anaphylactic degranulation. *Cytometry B Clin Cytom* 2012; 82:132-8.
50. Gibbs BF, Sabato V, Bridts CH, Ebo DG, Ben-Zimra M, Levi-Schaffer F. Expressions and inhibitory functions of CD300a receptors on purified human basophils. *Exp Dermatol* 2012; 21:884-6.
51. Sabato V, Boita M, Shubber S, Bridts CH, Shibuya A, De Clerck LS, et al. Mechanism of phosphatidylserine inhibition of IgE/FcepsilonRI-dependent anaphylactic human basophil degranulation via CD300a. *J Allergy Clin Immunol* 2014; 134:734-7.e3.
52. Ebo DG, Dombrecht EJ, Bridts CH, Aerts NE, de Clerck LS, Stevens WJ. Combined analysis of intracellular signalling and immunophenotype of human peripheral blood basophils by flow cytometry: a proof of concept. *Clin Exp Allergy* 2007; 37:1668-75.
53. Aerts NE, Dombrecht EJ, Bridts CH, Hagendorens MM, de Clerck LS, Stevens WJ, et al. Simultaneous flow cytometric detection of basophil activation marker CD63 and intracellular

- phosphorylated p38 mitogen-activated protein kinase in birch pollen allergy. *Cytometry B Clin Cytom* 2009; 76:8-17.
54. Verweij MM, De Knop KJ, Bridts CH, De Clerck LS, Stevens WJ, Ebo DG. P38 mitogen-activated protein kinase signal transduction in the diagnosis and follow up of immunotherapy of wasp venom allergy. *Cytometry B Clin Cytom* 2010; 78:302-7.
 55. Verweij MM, Sabato V, Nullens S, Bridts CH, De Clerck LS, Stevens WJ, et al. STAT5 in human basophils: IL-3 is required for its FcepsilonRI-mediated phosphorylation. *Cytometry B Clin Cytom* 2012; 82:101-6.
 56. Ebo DG, Bridts CH, Mertens CH, Hagendorens MM, Stevens WJ, De Clerck LS. Analyzing histamine release by flow cytometry (HistaFlow): a novel instrument to study the degranulation patterns of basophils. *J Immunol Methods* 2012; 375:30-8.
 57. Nullens S, Sabato V, Faber M, Leysen J, Bridts CH, De Clerck LS, et al. Basophilic histamine content and release during venom immunotherapy: insights by flow cytometry. *Cytometry B Clin Cytom* 2013; 84:173-8.
 58. Bridts CH, Sabato V, Mertens C, Hagendorens MM, De Clerck LS, Ebo DG. Flow cytometric allergy diagnosis: basophil activation techniques. *Methods Mol Biol* 2014; 1192:147-59.
 59. Cop N, Uyttebroek AP, Sabato V, Bridts CH, De Clerck LS, Ebo DG. Flow cytometric analysis of drug-Induced basophil histamine release. *Cytometry B Clin Cytom* 2016; 90:285-8.
 60. Joulia R, Mailhol C, Valitutti S, Didier A, Espinosa E. Direct monitoring of basophil degranulation by using avidin-based probes. *J Allergy Clin Immunol* 2017; 140:1159-62.e6.
 61. Mukai K, Chinthrajah RS, Nadeau KC, Tsai M, Gaudenzio N, Galli SJ. A new fluorescent-avidin-based method for quantifying basophil activation in whole blood. *J Allergy Clin Immunol* 2017; 140:1202-6.e3.
 62. Ebo DG, Elst J, van Houdt M, Pintelon I, Timmermans JP, Horiuchi T, et al. Flow cytometric basophil activation tests: Staining of exteriorized basophil granule matrix by fluorescent avidin versus appearance of CD63. *Cytometry B Clin Cytom* 2020. doi:10.1002/cyto.b.21868.
 63. Wedi B, Gehring M, Kapp A. The pseudoallergen receptor MRGPRX2 on peripheral blood basophils and eosinophils: Expression and function. *Allergy*. 2020 Sep;75:2229-2242. doi: 10.1111/all.14213. Epub 2020 Feb 15. PMID: 32003863. Comments in: Surface expression of MRGPRX2 on resting basophils: An area of controversy. Sabato V, Elst J, Van Houdt M, Bridts C, Mertens C, Ebo DG. *Allergy*. 2020 Sep;75:2421-2422. doi: 10.1111/all.14252. PMID: 32929729 and Reply to Sabato V et al "Surface expression of MRGPRX2 expression on resting basophils: An area of controversy". Wedi B, Gehring M, Kapp A. *Allergy*. 2020 Sep;75:2424-2427. doi: 10.1111/all.14446.
 64. Elst J, Sabato V, Hagendorens MM, van Houdt M, Faber MA, Bridts CH, et al. Basophil Activation Techniques: Staining of Exteriorized Granule Matrix. *Methods Mol Biol* 2020; 2163:213-8.
 65. Devouassoux G, Foster B, Scott LM, Metcalfe DD, Prussin C. Frequency and characterization of antigen-specific IL-4- and IL-13- producing basophils and T cells in peripheral blood of healthy and asthmatic subjects. *J Allergy Clin Immunol* 1999; 104:811-9.
 66. Ocmant A, Michils A, Schandene L, Peignois Y, Goldman M, Stordeur P. IL-4 and IL-13 mRNA real-time PCR quantification on whole blood to assess allergic response. *Cytokine* 2005; 31:375-81.
 67. Chirumbolo S. Commentary: The Expression of CD123 Can Decrease with Basophil Activation: Implications for the Gating Strategy of the Basophil Activation Test. *Front Immunol* 2016; 7:260.
 68. Cozon G, Ferrandiz J, Peyramond D, Brunet J. Detection of activated basophils using flow cytometry for diagnosis in atopic patients. *Allergol Immunopathol (Madr)* 1999; 27:182-7.
 69. Paris-Kohler A, Demoly P, Persi L, Lebel B, Bousquet J, Arnoux B. In vitro diagnosis of cypress pollen allergy by using cytofluorimetric analysis of basophils (Basotest). *J Allergy Clin Immunol* 2000; 105:339-45.


70. Sanz ML, Sanchez G, Gamboa PM, Vila L, Uasuf C, Chazot M, et al. Allergen-induced basophil activation: CD63 cell expression detected by flow cytometry in patients allergic to *Dermatophagoides pteronyssinus* and *Lolium perenne*. *Clin Exp Allergy* 2001; 31:1007-13.
71. Ocmant A, Peignois Y, Mulier S, Hanssens L, Michils A, Schandene L. Flow cytometry for basophil activation markers: the measurement of CD203c up-regulation is as reliable as CD63 expression in the diagnosis of cat allergy. *J Immunol Methods* 2007; 320:40-8.
72. Ozdemir SK, Guloglu D, Sin BA, Elhan AH, Ikinciogullari A, Misirligil Z. Reliability of basophil activation test using CD203c expression in diagnosis of pollen allergy. *Am J Rhinol Allergy* 2011; 25:e225-31.
73. Moneret-Vautrin DA, Sainte-Laudy J, Kanny G, Fremont S. Human basophil activation measured by CD63 expression and LTC4 release in IgE-mediated food allergy. *Ann Allergy Asthma Immunol* 1999; 82:33-40.
74. Erdmann SM, Heussen N, Moll-Slodowy S, Merk HF, Sachs B. CD63 expression on basophils as a tool for the diagnosis of pollen-associated food allergy: sensitivity and specificity. *Clin Exp Allergy* 2003; 33:607-14.
75. Ebo DG, Hagendorens MM, Bridts CH, Schuerwegh AJ, De Clerck LS, Stevens WJ. Flow cytometric analysis of in vitro activated basophils, specific IgE and skin tests in the diagnosis of pollen-associated food allergy. *Cytometry B Clin Cytom* 2005; 64:28-33.
76. Erdmann SM, Sachs B, Schmidt A, Merk HF, Scheiner O, Moll-Slodowy S, et al. In vitro analysis of birch-pollen-associated food allergy by use of recombinant allergens in the basophil activation test. *Int Arch Allergy Immunol* 2005; 136:230-8.
77. Ebo DG, Bridts CH, Hagendorens MM, De Clerck LS, Stevens WJ. Scampi allergy: from fancy name-giving to correct diagnosis. *J Investig Allergol Clin Immunol* 2008; 18:228-30.
78. Ocmant A, Mulier S, Hanssens L, Goldman M, Casimir G, Mascart F, et al. Basophil activation tests for the diagnosis of food allergy in children. *Clin Exp Allergy* 2009; 39:1234-45.
79. Javaloyes G, Goikoetxea MJ, Garcia Nunez I, Sanz ML, Blanca M, Scheurer S, et al. Performance of different in vitro techniques in the molecular diagnosis of peanut allergy. *J Investig Allergol Clin Immunol* 2012; 22:508-13.
80. Appel MY, Nachshon L, Elizur A, Levy MB, Katz Y, Goldberg MR. Evaluation of the basophil activation test and skin prick testing for the diagnosis of sesame food allergy. *Clin Exp Allergy* 2018; 48:1025-34.
81. Glaumann S, Nopp A, Johansson SG, Rudengren M, Borres MP, Nilsson C. Basophil allergen threshold sensitivity, CD-sens, IgE-sensitization and DBPCFC in peanut-sensitized children. *Allergy* 2012; 67:242-7.
82. Santos AF, Douiri A, Becares N, Wu SY, Stephens A, Radulovic S, et al. Basophil activation test discriminates between allergy and tolerance in peanut-sensitized children. *J Allergy Clin Immunol* 2014; 134:645-52.
83. Ruinemans-Koerts J, Schmidt-Hieltjes Y, Jansen A, Savelkoul HFJ, Plaisier A, van Setten P. The Basophil Activation Test reduces the need for a food challenge test in children suspected of IgE-mediated cow's milk allergy. *Clin Exp Allergy* 2019; 49:350-6.
84. Decuyper, II, Pascal M, Van Gasse AL, Mertens C, Diaz-Perales A, Araujo G, et al. Performance of basophil activation test and specific IgG4 as diagnostic tools in nonspecific lipid transfer protein allergy: Antwerp-Barcelona comparison. *Allergy* 2019.
85. Faber M, Sabato V, De Witte L, Van Gasse A, Hagendorens MM, Leysen J, et al. State of the art and perspectives in food allergy (part I): diagnosis. *Curr Pharm Des* 2014; 20:954-63.
86. Sainte-Laudy J, Sabbah A, Drouet M, Lauret MG, Loiry M. Diagnosis of venom allergy by flow cytometry. Correlation with clinical history, skin tests, specific IgE, histamine and leukotriene C4 release. *Clin Exp Allergy* 2000; 30:1166-71.
87. Platz IJ, Binder M, Marxer A, Lischka G, Valent P, Buhning HJ. Hymenoptera-venom-induced upregulation of the basophil activation marker ecto-nucleotide pyrophosphatase/phosphodiesterase 3 in sensitized individuals. *Int Arch Allergy Immunol* 2001; 126:335-42.

88. Binder M, Fierlbeck G, King T, Valent P, Buhning HJ. Individual hymenoptera venom compounds induce upregulation of the basophil activation marker ectonucleotide pyrophosphatase/phosphodiesterase 3 (CD203c) in sensitized patients. *Int Arch Allergy Immunol* 2002; 129:160-8.
89. Lambert C, Guilloux L, Dzviga C, Gourgaud-Massias C, Genin C. Flow cytometry versus histamine release analysis of in vitro basophil degranulation in allergy to Hymenoptera venom. *Cytometry B Clin Cytom* 2003; 52:13-9.
90. Sturm GJ, Bohm E, Trummer M, Weiglhofer I, Heinemann A, Aberer W. The CD63 basophil activation test in Hymenoptera venom allergy: a prospective study. *Allergy* 2004; 59:1110-7.
91. Erdmann SM, Sachs B, Kwiecien R, Moll-Slowsky S, Sauer I, Merk HF. The basophil activation test in wasp venom allergy: sensitivity, specificity and monitoring specific immunotherapy. *Allergy* 2004; 59:1102-9.
92. Eberlein-König B, Schmidt-Leidescher C, Rakoski J, Behrendt H, Ring J. In vitro basophil activation using CD63 expression in patients with bee and wasp venom allergy. *J Invest Allergol Clin Immunol* 2006; 16:5-10.
93. Eberlein-König B, Varga R, Mempel M, Darsow U, Behrendt H, Ring J. Comparison of basophil activation tests using CD63 or CD203c expression in patients with insect venom allergy. *Allergy* 2006; 61:1084-5.
94. Ebo DG, Hagendorens MM, Schuerwegh AJ, Beirens LM, Bridts CH, De Clerck LS, et al. Flow-assisted quantification of in vitro activated basophils in the diagnosis of wasp venom allergy and follow-up of wasp venom immunotherapy. *Cytometry B Clin Cytom* 2007; 72:196-203.
95. Scherer K, Weber JM, Jermann TM, Krautheim A, Tas E, Ueberschlag EV, et al. Cellular in vitro assays in the diagnosis of Hymenoptera venom allergy. *Int Arch Allergy Immunol* 2008; 146:122-32.
96. Peternelj A, Silar M, Bajrovic N, Adamic K, Music E, Kosnik M, et al. Diagnostic value of the basophil activation test in evaluating Hymenoptera venom sensitization. *Wien Klin Wochenschr* 2009; 121:344-8.
97. Ott H, Tenbrock K, Baron J, Merk H, Lehmann S. Basophil activation test for the diagnosis of hymenoptera venom allergy in childhood: a pilot study. *Klin Padiatr* 2011; 223:27-32.
98. Bonadonna P, Zanotti R, Melioli G, Antonini F, Romano I, Lenzi L, et al. The role of basophil activation test in special populations with mastocytosis and reactions to hymenoptera sting. *Allergy* 2012; 67:962-5.
99. Rietveld MJ, Schreurs MW, Gerth van Wijk R, van Daele PL, Hermans MA. The Basophil Activation Test Is Not a Useful Screening Tool for Hymenoptera Venom-Related Anaphylaxis in Patients with Systemic Mastocytosis. *Int Arch Allergy Immunol* 2016; 169:125-9.
100. Ebo DG, Lechkar B, Schuerwegh AJ, Bridts CH, De Clerck LS, Stevens WJ. Validation of a two-color flow cytometric assay detecting in vitro basophil activation for the diagnosis of IgE-mediated natural rubber latex allergy. *Allergy* 2002; 57:706-12.
101. Sanz ML, Gamboa PM, Garcia-Aviles C, Vila L, Dieguez I, Antepara I, et al. Flow-cytometric cellular allergen stimulation test in latex allergy. *Int Arch Allergy Immunol* 2003; 130:33-9.
102. Boumiza R, Monneret G, Forissier MF, Savoye J, Gutowski MC, Powell WS, et al. Marked improvement of the basophil activation test by detecting CD203c instead of CD63. *Clin Exp Allergy* 2003; 33:259-65.
103. Nettis E, Colanardi MC, Dambra PP, Capuzzimati L, Loria MP, Ferrannini A, et al. Flow cytometric basophil activation test: detection of CD63 expression as a useful aid to diagnosis of latex allergy. *Ann Allergy Asthma Immunol* 2006; 97:715-6.
104. Hemery ML, Arnoux B, Dhivert-Donnadieu H, Rongier M, Barbotte E, Verdier R, et al. Confirmation of the diagnosis of natural rubber latex allergy by the Basotest method. *Int Arch Allergy Immunol* 2005; 136:53-7.
105. Sanz ML, Garcia-Aviles MC, Tabar AI, Anda M, Garcia BE, Barber D, et al. Basophil Activation Test and specific IgE measurements using a panel of recombinant natural rubber latex

- allergens to determine the latex allergen sensitization profile in children. *Pediatr Allergy Immunol* 2006; 17:148-56.
106. Decuyper, II, Mangodt EA, Van Gasse AL, Claesen K, Uyttendaele A, Faber M, et al. In Vitro Diagnosis of Immediate Drug Hypersensitivity Anno 2017: Potentials and Limitations. *Drugs R D* 2017; 17:265-78.
 107. Sanz ML, Gamboa PM, Antepara I, Uasuf C, Vila L, Garcia-Aviles C, et al. Flow cytometric basophil activation test by detection of CD63 expression in patients with immediate-type reactions to betalactam antibiotics. *Clin Exp Allergy* 2002; 32:277-86.
 108. Gamboa PM, Garcia-Aviles MC, Urrutia I, Antepara I, Esparza R, Sanz ML. Basophil activation and sulfidoleukotriene production in patients with immediate allergy to betalactam antibiotics and negative skin tests. *J Investig Allergol Clin Immunol* 2004; 14:278-83.
 109. Torres MJ, Padial A, Mayorga C, Fernandez T, Sanchez-Sabate E, Cornejo-Garcia JA, et al. The diagnostic interpretation of basophil activation test in immediate allergic reactions to betalactams. *Clin Exp Allergy* 2004; 34:1768-75.
 110. Abuaf N, Rostane H, Rajoely B, Gaouar H, Autegarden JE, Leynadier F, et al. Comparison of two basophil activation markers CD63 and CD203c in the diagnosis of amoxicillin allergy. *Clin Exp Allergy* 2008; 38:921-8.
 111. De Week AL, Sanz ML, Gamboa PM, Aberer W, Sturm G, Bilo MB, et al. Diagnosis of immediate-type beta-lactam allergy in vitro by flow-cytometric basophil activation test and sulfidoleukotriene production: a multicenter study. *J Investig Allergol Clin Immunol* 2009; 19:91-109.
 112. Garcia-Ortega P, Marin A. Usefulness of the basophil activation test (BAT) in the diagnosis of life-threatening drug anaphylaxis. *Allergy* 2010; 65:1204.
 113. Torres MJ, Ariza A, Fernandez J, Moreno E, Laguna JJ, Montanez MI, et al. Role of minor determinants of amoxicillin in the diagnosis of immediate allergic reactions to amoxicillin. *Allergy* 2010; 65:590-6.
 114. Torres MJ, Romano A, Blanca-Lopez N, Dona I, Canto G, Ariza A, et al. Immunoglobulin E-mediated hypersensitivity to amoxicillin: in vivo and in vitro comparative studies between an injectable therapeutic compound and a new commercial compound. *Clin Exp Allergy* 2011; 41:1595-601.
 115. Salas M, Fernandez-Santamaria R, Mayorga C, Barrionuevo E, Ariza A, Posadas T, et al. Use of the Basophil Activation Test May Reduce the Need for Drug Provocation in Amoxicillin-Clavulanic Allergy. *J Allergy Clin Immunol Pract* 2018; 6:1010-8.e2.
 116. Uyttendaele AP, Sabato V, Cop N, Decuyper, II, Faber MA, Bridts CH, et al. Diagnosing cefazolin hypersensitivity: Lessons from dual-labeling flow cytometry. *J Allergy Clin Immunol Pract* 2016; 4:1243-5.
 117. Dreborg S. Methodological cutoff of basophil activation test and basophil activation test diagnostic value. *J Allergy Clin Immunol Pract* 2018; 6:1089-90.
 118. Mayorga C, Doña I, Perez-Inestrosa E, Fernández TD, Torres MJ. The Value of In Vitro Tests to Diminish Drug Challenges. *Int J Mol Sci* 2017; 18.
 119. Abuaf N, Rajoely B, Ghazouani E, Levy DA, Pecquet C, Chabane H, et al. Validation of a flow cytometric assay detecting in vitro basophil activation for the diagnosis of muscle relaxant allergy. *J Allergy Clin Immunol* 1999; 104:411-8.
 120. Sudheer PS, Hall JE, Read GF, Rowbottom AW, Williams PE. Flow cytometric investigation of peri-anaesthetic anaphylaxis using CD63 and CD203c. *Anaesthesia* 2005; 60:251-6.
 121. Kvedariene V, Kamey S, Ryckwaert Y, Rongier M, Bousquet J, Demoly P, et al. Diagnosis of neuromuscular blocking agent hypersensitivity reactions using cytofluorimetric analysis of basophils. *Allergy* 2006; 61:311-5.
 122. Sainte-Laudry J, Orsel I. Interest of a new flow cytometric protocol applied to diagnose and prevention of perianesthetic accidents by neuromuscular blocking agents. *Rev Fr Allergol* 2008; 48:5.

123. Hagau N, Gherman-Ionica N, Sfichi M, Petrisor C. Threshold for basophil activation test positivity in neuromuscular blocking agents hypersensitivity reactions. *Allergy Asthma Clin Immunol* 2013; 9:42.
124. Uyttebroek AP, Sabato V, Leysen J, Bridts CH, De Clerck LS, Ebo DG. Flowcytometric diagnosis of atracurium-induced anaphylaxis. *Allergy* 2014; 69:1324-32.
125. Li J, Best OG, Rose MA, Green SL, Fulton RB, Fernando SL. Integrating basophil activation tests into evaluation of perioperative anaphylaxis to neuromuscular blocking agents. *Br J Anaesth* 2019; 123:e135-e43.
126. Van Gasse AL, Elst J, Bridts CH, Mertens C, Faber M, Hagendorens MM, et al. Rocuronium Hypersensitivity: Does Off-Target Occupation of the MRGPRX2 Receptor Play a Role? *J Allergy Clin Immunol Pract* 2019; 7:998-1003.
127. Seitz CS, Brouckebus EB, Trautmann A. Diagnostic testing in suspected fluoroquinolone hypersensitivity. *Clin Exp Allergy* 2009; 39:1738-45.
128. Lobera T, Audicana MT, Alarcon E, Longo N, Navarro B, Munoz D. Allergy to quinolones: low cross-reactivity to levofloxacin. *J Investig Allergol Clin Immunol* 2010; 20:607-11.
129. Ben Said B, Berard F, Bienvenu J, Nicolas JF, Rozieres A. Usefulness of basophil activation tests for the diagnosis of IgE-mediated allergy to quinolones. *Allergy* 2010; 65:535-6.
130. Aranda A, Mayorga C, Ariza A, Dona I, Rosado A, Blanca-Lopez N, et al. In vitro evaluation of IgE-mediated hypersensitivity reactions to quinolones. *Allergy* 2011; 66:247-54.
131. Rouzaire P, Nosbaum A, Denis L, Bienvenu F, Berard F, Cozon G, et al. Negativity of the basophil activation test in quinolone hypersensitivity: a breakthrough for provocation test decision-making. *Int Arch Allergy Immunol* 2012; 157:299-302.
132. Van Gasse AL, Sabato V, Uyttebroek AP, Elst J, Faber MA, Hagendorens MM, et al. Immediate moxifloxacin hypersensitivity: Is there more than currently meets the eye? *Allergy* 2017; 72:2039-43.

membrane markers




= Fluorochrome labelled avidin

signalosome

= phosphorylated p38MAPK


= phosphorylated STAT5



mediators (or mRNA thereof)

= fluorochrome labelled mRNA



= DAO-labelled histamine









= released histamine


intracellular calcium staining

= Fluo4AM- labelled Calcium







Allergen dose response curve





Allergen sensitivity and specificity curve at 1 $\mu\text{g/mL}$







Selection basophils




Buffer




Sesame 10 $\mu\text{g/mL}$




Sesame oil 0.16 $\mu\text{g/mL}$






















Sesame oil 1.6 $\mu\text{g/mL}$





Sesame oil 16 $\mu\text{g/mL}$








Selection basophils



AIBAT: negative



AIBAT: positive

